

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/15170/2024

Dated: 18.01.2025

NOTICE**Sub:-** Tender No.T547/ 2024 - Cancellation Notice – PCR Reagents & Kits - reg.**Ref:-** 1. Tender Inviting Notice published dated 31.08.2024

2. Technical bid opening held on 14.10.2024 @ 11.00 A M

3. Reports of Demonstration/ document evaluation of PCR Reagents & Kits Received on 04.01.2025

1. The reports of technical document evaluation/ demonstration of PCR Reagents & Kits as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

| SI No | Equipment Name | Selected Bidders |
|-------|---|---------------------------------------|
| II | Japanese Encephalitis / West Nile PCR | M/s Huwel Lifesciences Pvt. Ltd |
| | | M/s GCC Biotech India Private Limited |
| IV | Influenza A/B/H1N1 Multiplex PCR Kit | M/s GCC Biotech India Private Limited |
| V | Covid-19 RT-PCR Kit | M/s. Siemens Healthcare Pvt Ltd |
| | | M/s GCC Biotech India Private Limited |
| VII | Kyasanur Forest Disease PCR | M/s Huwel Lifesciences Pvt. Ltd |
| IX | Total Viral Nucleic Acid Extraction Kits (50 reaction kit) | M/s Q-line Biotech Private Limited |
| | | M/s Huwel Lifesciences Pvt. Ltd |
| | | M/s GCC Biotech India Private Limited |

2. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

II. Japanese Encephalitis / West Nile PCR

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

III. Dengue/Chikungunya/Zika Multiplex PCR Kit

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

C. M/s Huwel Lifesciences Pvt. Ltd

Reason for rejection: The kit failed to detect positive test samples.

D. M/s GCC Biotech India Private Limited

Reason for rejection: The kit failed to detect positive test samples correctly.

IV. Influenza A/B/H1N1 Multiplex PCR Kit

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

C. M/s Huwel Lifesciences Pvt. Ltd

Reason for rejection:

| Clause No | Technical specification | Deviations noted |
|-----------|--|------------------|
| III.2 | Should have extractable exogenous internal control | Not available |

V. Covid-19 RT-PCR Kit

A. M/s Q-line Biotech Private Limited

Reason for rejection: The reagents are stable only for 2 months after opening.

B. M/s Huwel Lifesciences Pvt. Ltd

Reason for rejection:

VI. Respiratory Viral Panel Multiplex PCR Kit

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

C. M/s Huwel Lifesciences Pvt. Ltd

Reason for rejection:

| | | |
|------|---|---|
| V 1. | Real time Multiplex PCR kit for detecting Inf A, Inf B, H1N1, Rhinovirus, Coronaviruses (NL63, 229E, OC43, HKU1), Parainfluenza (1,2,3,4), Human Metapneumovirus (A&B), Adenovirus, RSV(A &B), Enterovirus, Echovirus, and Mycoplasma pneumoniae with internal controls | No, Mycoplasma pneumoniae not available |
| | Comments | The kit failed to detect low positive test samples. |

D. M/s GCC Biotech India Private Limited

Reason for rejection: The kit failed to detect positive test samples correctly.

Vii. Kyasanur Forest Disease PCR

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

C. M/s GCC Biotech India Private Limited

Reason for rejection: Not Quoted

Viii. Encephalitis Panel - Viral Multiplex Panel

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

C. M/s Huwel Lifesciences Pvt. Ltd

Reason for rejection:

| Clause No | Technical specification | Deviations noted |
|-----------|---|---|
| VII.1 | Real time PCR Kit for detecting CMV, EBV, Adenovirus, Herpes simplex virus (1&2), Varicella zoster virus, enterovirus, Parechovirus, human herpes virus 6,7, Parvo B19 with internal controls | HHV 6 and HHV 7 not available |
| | Comments | The kit failed to detect positive test samples. |

D. M/s GCC Biotech India Private Limited

Reason for rejection: The kit failed to detect positive test samples.

IX. Total Viral Nucleic Acid Extraction Kits (50 reaction kit)

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

III. Covid-19 Rapid Antigen Test

A. M/s. Siemens Healthcare Pvt Ltd

Reason for rejection: Not Quoted

B. M/s Q-line Biotech Private Limited

Reason for rejection: Not Quoted

C. M/s Huwel Lifesciences Pvt. Ltd

Reason for rejection: Not Quoted

D. M/s GCC Biotech India Private Limited

Reason for rejection: Not Quoted

3. Due to administrative reason the below mentioned items were cancelled.

| SI No | Item Description |
|-------|--|
| I | PCR Test Kits |
| II | Dengue/Chikungunya/Zika Multiplex PCR Kit |
| IV | Influenza A/B/H1N1 Multiplex PCR Kit |
| VI | Respiratory Viral Panel Multiplex PCR Kit |
| VII | Kyasanur Forest Disease PCR |
| VIII | Encephalitis Panel - Viral Multiplex Panel |
| X | Covid-19 Rapid Antigen Test |

4. The online price bid of the eligible tenderer mentioned above will be opened on 21.01.2025 @ 11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager