

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/216/2025

Dated: 24.10.2025

NOTICE

Sub:- Tender No.T563/ 2025 - Results of demonstration/ document evaluation & date of price bid opening – various equipment- reg:-

- Ref:-**
1. Tender Inviting Notice published dated 07.01.2025
 2. Technical bid opening held on 05.02.2025 & 12.05.2025 @ 11.00 AM
 3. Eligible list published on 08.06.2025
 4. Demonstration/ document evaluation of various equipment conducted on 27.08.2025

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl.No	Equipment	Selected Bidders	Selected Model
1.	Elisa Reader and Washer	M/s.UNIVERSAL AGENCIES	InviRead Touch & InviWash Touch/ Invitronix
2.	Plasma Expressor	M/s.Scientific Medical Equipments (Kerala) Pvt.Ltd	PE-I/ ELECTRORAFT
		M/s.Haemolab Equipments	Labtop/ Indian
		M/s.KOCHIN INLAB EQUIPMENTS INDIA PVT LTD	M1.0/TERUMO PENPOL

II. the other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T563/43/2025: Elisa Reader and Washer**1. M/s.ASPEN DIAGNOSTICS PVT LTD**

Rejected model: Lisa plus, Lisa wash/Aspen Biomedical

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
I.22	Software capable of entering patient ids, programing of tests with multiple standards, standard curve fit methods, storage and reuse of calibration curves, test results and full bidirectional control of the reader, required computer and laser printer should be provided.	This feature is not available
II.12	Provision for connecting minimum two wash bottles and selecting as required	This feature is not available
II.13	Should have cross wise aspiration	This feature is not available

28.10.2025

IV. The online price bid of the eligible tenderers mentioned above will be opened on ~~09.2025~~ 28.10.2025 @ 11.AM at the etender portal.

Dr Arun P V
General Manager