



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/1695/2025

Dated: 02.12.2025

NOTICE

Sub:- Tender No.T567/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment- reg.

- Ref:-**
1. Tender Inviting Notice published dated 05.02.2025
 2. Technical bid opening held on 08.05.2025 @ 11.00 A M
 3. Eligible list for demonstration published dated 03.06.2025
 4. Demonstation/ document evaluation of various equipment is received on 23.05.2025

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1	OT Light Shadow-less Ceiling Mount Single Dome	M/s. Kerala Surgical Equipment Co	Penta Lumeo 6000/Palakkad Surgical
2	OT Light Shadow-less Ceiling Mount Double Dome	M/s.Lakshmi Surgicals	Karl Storz /Karl Storz

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. KMSCL/EP/T567/90/2025: OT Light Shadow-less Ceiling Mount Single Dome

A. M/s. STAAAN BIO-MED ENGINEERING PRIVATE LIMITED

Rejected model: DISCOVER/STAAAN BIO.ME

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
17	Colour rendering index should be 95 or more	Not meeting the specification

B. M/s. HARDIK MEDI TECH

Rejected model: ALFA 12/ Surgident (India)

Reason for rejection; Sample not produced

C. M/s. Dynamic International

Rejected model: D-NANO/DYNAMIC

Reason for rejection; Sample not provided

D. M/s.BIOCARE DIAGNOSTICS

Rejected model; ECO PLUS 140/ENERTECH

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
17	Colour rendering index should be 95 or more	Not meeting the specification

2. KMSCL/EP/T567/89/2025: OT Light Shadow-less Ceiling Mount Double Dome

A. Dynamic International

Rejected model: DHI-96D/DYNAMIC

Reason for rejection: Sample not produced

B. HARDIK MEDI TECH

Rejected model: ALFA 12+12/Surgident (India)

Reason for rejection; Sample not produced

C. M/s.Kerala Surgical Equipment Co

Rejected model: Spectra Lumeo/Palakkad Surgical

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
8	Colour rendering index should be 95 or more	Not meeting the specification

D. M/s.BIOCARE DIAGNOSTICS

Rejected model: ECO PLUS 140+140/ENERTECH

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
8	Colour rendering index should be 95 or more	Not meeting the specification

E. M/s.ANTS LIFECARE PRIVATE LIMITED

Rejected model: ILUX 10-10 /SHALYA XcelLance

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	Shadow level not acceptable limit

F. M/s.STAAN BIO-MED ENGINEERING PRIVATE LIMITED

Rejected model: DISCOVER/STAANBIO-MED

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Minimum light field size should be adjustable and range from 18 - 30 cm or better	Not meeting the specification

08/12/2025

III. The online price bid of the eligible tenderer mentioned above will be opened on 08/12/2025 @ 11.AM at the e-tender portal.

Dr Arun P V

General Manager

