



#### KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking) Thycaud.P.O., Thiruvananthapuram-14. Tele Fax No. 0471-2945647 0471-2945646, 0471-2945600

E-mail id: eptenders.kmscl@kerala.gov.in CIN: U24233KL2007SGC021616 GSTIN: 32AADCK4029M1ZK

www.kmscl.kerala.gov.in

Dated: 13.11.2025

### **EP Division**

No: KMSCL/EP/7791/2025

### NOTICE

- Sub:- Tender No.T574(R)/ 2025 Results of demonstration/ document evaluation & Date of pricebid opening various equipment -reg;-
- Ref:- 1. Tender Inviting Notice published dated 14.05.2025
  - 2. Technical bid opening held on 09.06.2025, 19.06.2025, 28.06.2025 @ 11.00 AM
  - 3. Eligible list published on 24.09.2025
  - 4. Demonstation/ document evaluation of various equipment conducted on 03.11.2025
- I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Wheel Chair	M/s DESINTOX TECHNOLOGIES	5010/ MOTHERMED

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

# 1. KMSCL/EP/T574(R)/551C/2025 Wheel Chair

A. M/s Willson Medical System Pvt. Ltd. Rejected model: WM-8504/Willson Medica Reasons for rejection: Sample not produced

B. M/s.Dynamic International

Rejected model; DWC-1174/DYNAMIC Reason for rejection: Sample not produced

## 2. KMSCL/EP/T574(R)/277/2025Alpha Bed

A. M/s. Kerala Surgical Equipment Co Rejected model: AMT 053/ Life Line

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
9	Cycle time of inflation & deflation should be 3/5 minutes.	Not Meeting the specification
10	Pump should have visual low pressure indicator / alarm.	Not Meeting the specification

IV. The online price bid of the eligible tenderer mentioned above will be opened on ₹.₽₽.2025 @ 11.AM at the e-tender portal.

Dr Arun P V General Manager