



KERALA MEDICAL SERVICES CORPORATION LTD.,
(Dept. of Health & Family Welfare, Govt. of Kerala)
Thycaud P.O., Thiruvananthapuram-14
Tele Fax No : 0471-2945647,
0471-2945600
Email id : edrugs.kmscl@kerala.gov.in
CIN : U24233KL2007SGC021616

Essential Drugs Division
No. KMSCL/ED/7-RT/2025

www.kmscl.kerala.gov.in
Date : 22.08.2025

NOTICE

(Drugs/Supplies – Category III-VIII)

Sub:- KMSCL – Re -tender for the Procurement of Drugs/Supplies for the year 2025-26 –Technical Bid Evaluation – List of selected/rejected bidders/products for Price Bid Evaluation - Published – reg.

- Ref :-**
1. E-tender No. KMSCL/ED/III-VIII/RT/2025/017 Dated 03.07.2025.
 2. This office notice even dated 31.07.2025

The Technical bids submitted in response to the e-tender referred 1st above for the procurement of Drugs/Supplies for the year 2025-26 along with the clarifications/complaints received in respect of the notice 2nd cited were scrutinized. The list of selected/rejected bidders/products based on the scrutiny of the documents/complaints is published below.

The price bids of all the qualified bidders will be opened and evaluated from **2.00 Pm onwards on 23.08.2025.**

Sl.No	Drug Code	Drug Name	Strength	Unit	Decision Taken
Bid No. 1		M/s. Newtech Medical Devices Pvt. Ltd., Haryana			
1.	S27256	DISPOSABLE CVP CATHETER TRIPLE LUMEN (SELDINGER TECHNIQUE)	7F	1 No	<ol style="list-style-type: none">1. The balance EMD (offline) by way of DD/BG as per clause 3.2.4 is not submitted.2. Notary attested copy of Valid Market Standing Certificate and Previous Market Standing Certificate to prove three year Market Standing not submitted for the quoted item.3. Notary Attested copy

Sl.No	Drug Code	Drug Name	Strength	Unit	Decision Taken
					<p>of valid Non Conviction Certificate against MD License (MFG/MD/ 2023/000475) not submitted.</p> <p>4. Notary attested copy of Power of Attorney in non-judicial stamp paper/Resolution of Board as per Clause no. 6.5.5 not submitted.</p> <p>5. Notary attested copy of Performance statement in Annexure IV not submitted.</p> <p>6. The license no (MFG/MD/ 2023/000475) is not mentioned in the Details of the bidder and Manufacturing unit in Annexure VII submitted. Hence offer of the firm is rejected.</p>
Bid No. 2		M/s. Poly Medicure Ltd., New Delhi			
1.	S27256	DISPOSABLE CVP CATHETER TRIPLE LUMEN (SELDINGER TECHNIQUE)	7F	1 No	The offer of all the products is accepted.
2.	S27219	ECG ELECTRODES	Null	1 No	
Bid No. 3		M/s. Medsource Ozone Biomedical Pvt Ltd , Haryana			
1.	D26026	DENGUE LATEX AGGLUTINATION TEST KIT (RAPID METHOD)	1 Test	1 No	<p>1. Notary attested copy of Valid Market Standing Certificate not submitted for the following items;</p> <p>D26023-HEPATITIS B SURFACE ANTIGEN</p>

Sl.No	Drug Code	Drug Name	Strength	Unit	Decision Taken
2.	D26023	HEPATITIS B SURFACE ANTIGEN SCREENING KIT-EIA VISUAL ASSAY TEST	1 Test	1 No	SCREENING KIT-EIA VISUAL ASSAY TEST-1 Test
3.	D26019	HIV 1 & 2 ANTIBODY SCREENING EIA VISUAL ASSAY (RAPID METHOD)	1 Test	1 No	D26019-HIV 1 & 2 ANTIBODY SCREENING EIA VISUAL ASSAY (RAPID METHOD)-1 Test
4.	D26018	HIV 1&2 ANTIBODY SCREENING-ELISA TEST KIT	1 Test	1 No	<p>2. The samples submitted for the following items for qualitative evaluation as per clause 6.16 of the tender document is rejected by the technical committee due to following reasons.</p> <p>D26023 - HEPATITIS B SURFACE ANTIGEN SCREENING KIT-EIA VISUAL ASSAY TEST (sensitivity poor in 3 centres .False negatives are obtained in 3 centres.)</p> <p>D26019 -HIV 1 & 2 ANTIBODY SCREENING EIA VISUAL ASSAY (RAPID METHOD) (sensitivity poor in all centres , weak positive bands and false negatives.)</p> <p>D26018- HIV 1&2 ANTIBODY SCREENING-ELISA TEST KIT (poor sensitivity in 3 centres)</p> <p>Hence offer of the products is rejected.</p> <p>The offer of the other product is accepted.</p>

Sl.No	Drug Code	Drug Name	Strength	Unit	Decision Taken
Bid No. 4		M/s. Surgimed Devices , Ahmedabad			
1.	S27250	SICS BLADE (CRESANT)	2.5mm/2.6m m angled bevel up	1 No	<p>1. Bid Offer Form in Annexure II (Pdf & Excel) not submitted.</p> <p>2. The online EMD of Rs.50,000/- and balance EMD (offline) by way of DD/BG as per clause 3.1.3 & 3.2.4 is not submitted.</p> <p>3. Notary attested copy of Average Annual Turnover statement for last three financial years in Annexure III is not submitted.</p> <p>4. Notary attested copy of the following documents is not submitted for the quoted items;</p> <p>a) MD License with Product Permit.</p> <p>b) Valid Market Standing Certificate & Previous Market Standing Certificate issued by the licensing Authority to prove three years Market Standing.</p> <p>c) Non Conviction</p>
2.	S27251	SICS BLADE (KERATOME)	2.8mm angled bevel up	1 No	
3.	S27252	SICS BLADE (SIDE PORT)	15° STRAIGHT LANCE TIP	1 No	

Sl.No	Drug Code	Drug Name	Strength	Unit	Decision Taken
					<p>Certificate.</p> <p>d) Performance Statement in Annexure IV.</p> <p>e) Power of Attorney/ Resolution of Board.</p> <p>f) Undertaking/Declaration in Annexure V.</p> <p>g) Audited Balance Sheets and Profit & Loss Statement for the years 2021-22, 2022-23 & 2023-24</p> <p>h) Details of Technical Personnel employed in the manufacturing and testing unit.</p> <p>i) Details of the Bidders and Manufacturing units in Annexure.</p> <p>5. The samples for the quoted items required for the qualitative evaluation as per clause 6.16 of the tender document is not submitted.</p> <p>Hence offer of the firm is rejected.</p>
Bid No. 5		M/s.SD Biosensor Healthcare Pvt Ltd , Haryana			
1.	D26014	DENGUE IGM CAPTURE ELISA (1,2,3,4) KIT	1 Test	1 No	1. The balance EMD

Sl.No	Drug Code	Drug Name	Strength	Unit	Decision Taken
2.	D26026	DENGUE LATEX AGGLUTINATION TEST KIT(RAPID METHOD)	1 Test	1 No	<p>(offline) by way of DD/BG as per clause 3.2.4 is not submitted.</p> <p>2. The samples for the quoted items required for the qualitative evaluation as per clause 6.16 of the tender document is not submitted</p> <p>Hence offer of the firm is rejected.</p>
3.	D26019	HIV 1 & 2 ANTIBODY SCREENING EIA VISUAL ASSAY(RAPID METHOD)	1 Test	1 No	
Bid No. 6		M/s. Aurolab , Tamil Nadu			
1.	S27250	SICS BLADE (CRESANT)	2.5mm/2.6m m angled bevel up	1 No	The offer of all the products is accepted.
2.	S27251	SICS BLADE (KERATOME)	2.8mm angled bevel up	1 No	
3.	S27252	SICS BLADE (SIDE PORT)	15° STRAIGHT LANCE TIP	1 No	
Bid No. 7		M/s. Meril Diagnostics Pvt.Ltd, Gujarat			
1.	D26018	HIV 1&2 ANTIBODY SCREENING-ELISA TEST KIT	1 Test	1 No	The offer of the products is accepted.

Sd/-

MANAGING DIRECTOR
KERALA MEDICAL SERVICES CORPORATION LTD.
(TENDER INVITING AUTHORITY)