



## KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/19902/2024

Dated: 03.06.2025

### NOTICE

**Sub:-** Tender No.T556/ 2024 - Cancellation Notice – OT Light LED Dual Dome without Camera reg:-

**Ref:-** 1. Tender Inviting Notice published dated 14.11.2024

2. Technical bid opening held on 05.12.2024 @ 11.00 AM

3. Demonstration/ document evaluation of OT Light LED Dual Dome without Camera conducted on 08-05-2025

I. The reports of technical document evaluation/ demonstration of OT Light LED Dual Dome without Camera as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows:

A. M/s. ANTS LIFECARE PRIVATE LIMITED

Rejected model: SHALYA Ilux 10-10 OR Lights./SHALYA XceiLance

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1	Should be a Surgical Light unit incorporating the latest LED technology shadow less operating light field with the following specifications.	Not meeting the specification

B. M/s.Dynamic International

Rejected model: DHL-04/DYNAMIC INTERNATIONAL

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	No model number in the machine

C. M/s.Kerala Surgical Equipment Co

Rejected model; Penta Lumeo- DD 8080/Palakkad Surgical

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
13	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	No valid certificate

D. M/s.Sree Gokulam Healthcare Pvt. Ltd  
Rejected model: HyLED c5/c5/Nanjing Mindary Bio-Medical  
Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3.a	LUX intensity 1,40,000 Lux & Satellite 1,40,000 Lux or above.	128000 Lux only

E. M/s.STAAN BIO-MED ENGINEERING PRIVATE LIMITED  
Rejected model: DISCOVER/STAAN BIO-MED  
Reason for rejection:

Clause No	Technical Specification	Deviations Noted
13	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	No valid certificate

III. Due to technical reason the tender for **KMSCL/EP/T556/306A/2024: OT Light LED Dual Dome without Camera** is cancelled.

**Dr. Vinay Goyal IAS**  
Managing Director