

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division**www.kmscl.kerala.gov.in****No: KMSCL/EP/18237/2024****Dated: 07.05.2025****NOTICE**

Sub:- Tender No.T552(R)/ 2024 -Results of demonstration/ document evaluation & Date of price bid opening –
Penta Head Microscope- reg.

Ref:- 1. Tender Inviting Notice published dated 15.10.2024

2. Technical bid opening held on 11.11.2024 @ 11.00 A M

3. Reports of demonstration/ document evaluation of Penta Head Microscope received on 29.04.2025

I. The reports of technical document evaluation/ demonstration of Penta Head Microscope as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Penta Head Microscope	M/s.Radical Scientific Equipments Pvt. Ltd	RXLr-5 Code: NX/ Radical Scientific

II. The other tenderers participated in the tender are rejected due to the reason mentioned against their name below;

A. M/s. Adelta Optec

Rejected model; AV 5liT30 SF22 Product ordering code:AP42p/Adelta optec

Reason for rejection: Sample not produced

B. M/s. Alan Medical & Laboratory Products

Rejected model : KXL-3003/Kwality Micro Scientific

Reason for rejection:

Clause No	technical specification	Deviation noted
7	The pixel size of the camera shall be below 2 microns to provide best image resolution when working with low magnification objectives.	4.14 micrometer
19	Filter-blue pointer- green LED	Not demonstrated

C. M/s.Opto Electronics Inc

Rejected model: Nexus - 5 + 8 MP 4KHDMI Camera + Table/i7 Opto Electronics

Reason for rejection;

Clause No	technical specification	Deviation noted
7	The pixel size of the camera shall be below 2 microns to provide best image resolution	2x2 micrometer

III. The online price bid of the eligible tenderer mentioned above will be opened on 02.05.2025 @ 11.AM at the e-tender portal.

Dr. Shibulal A
General Manager