



## KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/1608/2025

Dated: 02.09.2025

### NOTICE

**Sub:-** Tender No.T565/ 2025 - Cancellation Notice – Automated Tissue Stainer - reg.

**Ref:-** 1. Tender Inviting Notice published dated 05.02.2025

2. Technical bid opening held on 01.03.2025 & 10.03.2025 @ 11.00 A M

3. Eligible list for equipment demonstration published on 03.06.2025

4. Reports of demonstration/ document evaluation of Automated Tissue Stainer received on 14.08.2025

I. The reports of technical document evaluation/ demonstration of Automated Tissue Stainer per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are tabulated as follows:

SI No	Name of the Equipment	Name of the selected bidders	Selected models
1.	Automated Tissue Stainer	M/s. Bions Medical Systems Pvt Ltd	Leica ST5010 Autostainer XL/Leica Bio systems

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s. UNIVERSAL AGENCIES

Rejected model: MYREVA SS-30/ESPECIALIDADES

Reason for rejection:

Clause No	Technical Specification	Deviations noted
2.	Should have the capability to stain H & E and Pap simultaneously.	Not meeting the specification

III. In accordance with procurement norms, the tender for KMSCL/EP/T565/1767/2025: Automated Tissue Stainer is cancelled.

**Dr Arun P V**

General Manager