



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No. I/378269/2026

Date:10-09-2026

NOTICE

Sub:- Tender No.T580/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment- reg.

Ref:- 1. Tender Inviting Notice published dated 24.10.2025

2. Technical bid opening held on 22.01.2026 @ 11.00 A M

3. Eligible list for demonstration published dated 02.06.2026

4.Demonstration/ document evaluation of various equipment is conducted on 24.06.2026

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	Digital Radiography System Type B	M/s. KTPL	GF 85/Samsung Electronics Co Ltd
2.	Digital Radiography System Type C	M/s. Allengers Medical Systems	MARS-50/ Allengers Medical Systems
3	Digital Radiography System Type D	M/s.Allengers Medical Systems	MARS-30/ Allengers Medical Systems

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T580/34B/2025,: Digital Radiography System Type B

1. M/s.Trivitron Healthcare Pvt Ltd

Rejected Model: Ultisys 82/Kiran Medical Systems

Reason for rejection:

Clause No	Technical Specification	Deviations Noted	B.
II. 6	Should have inbuilt camera for live monitoring of patients during positioning and	No	

	exposure (Optional)	
III.3	All movements should have electromagnetic brakes with fully counter balanced mechanism with collision protection sensors.	Not adequate
III.4	It should have tube head to display, image preview, patient demographics, x ray parameters adjustments, FFD/SID, tube angle, standard collimation size adjustment and worklist access, filter selection,	Image preview not demonstrated
IV.3	Should have tracking with X-Ray tube on Z and X axis.	Not adequate
11	Individual adjustment of collimation blades should be possible from the main console.	Not demonstrated
VIII.3	Automatic Source tilting method Image Stitching Software up to 3 or 4 images must be provided for performing full leg full spine / long body imaging in both vertical bucky and table bucky	Only image displayed

KMSCL/EP/T580/34C/2025: Digital Radiography System Type C

1. M/s.Trivitron Healthcare Pvt Ltd
 Rejected Model; Ultisys 52/Kiran Medical Systems
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
IV.1	Should be vertical column based X-ray tube with auto-tracking with detectors on Z axis and auto positioning as per user APR view selected.	No
V.3	Should have tracking with x ray tube on Z – axis and auto positioning as per user APR view selected.	no auto tracking

C.

KMSCL/EP/T580/34D/2025: Digital Radiography System Type D

1. M/s.Trivitron Healthcare Pvt Ltd
 Rejected Model: Ultisys 40/ Kiran Medical Systems
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	After sales service is not satisfactory

2. M/s.American Meditek Systems Pvt.Ltd
 Rejected Model: MedRay-30/ American Meditek
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	During the demonstration, the video recorded by the American Meditek system had very poor image quality.

III. The online price bid of the eligible tenderer mentioned above will be opened on
14.09.2026 @ 11.AM at the e-tender

Rakesh R U
GENERAL MANAGER