



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/7790/2025

Dated: 14.05.2026

NOTICE

Sub:- Tender No.T573/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – Delivery Kit-reg.

- Ref:-**
1. Tender Inviting Notice published dated 14.05.2025
 2. Technical bid opening held on 02.06.2025 @ 11.00 A M
 3. Eligible list for demonstration published dated 06.05.2026
 4. Re-Demonstration/ document evaluation of Delivery Kit is conducted on 16.02.2025

I. The reports of technical document evaluation/ demonstration of Delivery Kit as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	Normal Delivery Kits	M/s NEW CARE HYGIENE PRODUCTS	Normal Delivery Kits/Newcare
2.	Delivery Kits for C-Section	M/s.NEW CARE HYGIENE PRODUCTS	Delivery Kits For C Section/New care

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. NO. KMSCL/EP/T573/441A/2025: Normal Delivery Kits

- 1.M/s PLASTI SURGE INDUSTRIES PVT LTD
 Rejected model: PSI-365/Plasti Surge Industries
 Reason for rejection:

Clause No	Technical Specification	Deviations Noted
vii	Identification band for mother & baby , Ethylene-vinyl acetate (EVA)/vinyl material, 1.5 mm thickness – 1 each	Not good quality
xi	Absorbent Mop , 30 x 30 cms, eight layered gauze with tail– 1 Nos	fibres are widely woven
	Comments	During the clinical evaluation quality issues noted in gauze & plastic apron

2. M/s Trivandrum Surgicals and Lab Supplies
 Rejected model; NORMAL DELIVERY KIT/RAHA INDUSTRIAL
 Reason for rejection : Sample not produced

3. M/s Parkins Enterprises

Rejected model; NDKOI/ IMM Aculati

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
vii	Identification band for mother & baby, Ethylene-vinyl acetate (EVA)/vinyl material, 1.5 mm thickness – 1 each	Material not good quality
2	Packing: Item i to xi in a single sterile pack. Surgeon's gown is to be packed separately inside the kit.	Gown not kept in separate plastic sterile packing

B. No. KMSCL/EP/T573/441B/2025: Delivery Kits for C-Section

1. M/s. PLASTI SURGE INDUSTRIES PVT LTD

Rejected model; PSI-366/ Plasti Surge Industries

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	During the clinical evaluation, quality issues noted in gauze, Absorbent Mop & plastic apron

III. The online price bid of the eligible tenderer mentioned above will be opened on 05.05.2026 @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager