



# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/10407/2024

Dated: 05.04.2025

## NOTICE

**Sub:-** Tender No.T543/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – various items - reg.

- Ref:-**
1. Tender Inviting Notice published dated 11.06.2024
  2. Technical bid opening held on 03.07.2024 @ 11.00 A M
  3. Eligible list for demonstration published dated 18.12.2024
  4. Demonstation/ document evaluation of various items conducted on 06.03.2025

I. The reports of technical document evaluation/ demonstration of various items as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1.	Diathermy with Bipolar Cutting & Coagulation System	M/s. Divlabs Healthcare Pvt. Ltd.	Trinity QRF/Divlabs
2.	Surgeon Mobile Swivel Stool	M/s. Kerala Surgical Equipment co.	Medik- SGC/Medik Enterprises
		M/s. C CUBE ADVANCED TECHNOLOGIES	SRGST-100/C CUBE ADVANCED

II. The other tenders participated in the tender are rejected due to the reasons mentioned against their names below;

### A. KMSCL/EP/T543/1045/2024: Diathermy with Bipolar Cutting & Coagulation System

1. M/s. Lakshmi surgical

Rejected model: KLS MARTIN Diathermy with Bipolar cutting and coagulation system/KLS Martin

Reasons for rejection:

Clause No	Technical Specification	Deviations Noted
7	It should have durable metal foot pedal should with color coded independent foot pads	No foot switch for coagulation
8	It should have integrated irrigation system	No integrated irrigation system

III. Due to admonistrative reason the teder for **KMSCL/EP/T543/1045/2024: Diathermy with Bipolar Cutting & Coagulation System** is cancelled.

IV. The online price bid of the eligible tenderer mentioned above will be opened on ~~07~~04.2025 @ 11.AM at the etender portal.

**Dr. Shibulal A**  
General Manager