



# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/216/2025

Dated: 29.10.2025

## NOTICE

**Sub:-** Tender No.1563/ 2025 - Results of demonstration/ document evaluation & date of price bid opening –  
Vaccum Extractor- reg:-

- Ref:-**
1. Tender Inviting Notice published dated 05.03.2025
  2. Technical bid opening held on 26.03.2025 @ 11.00 AM
  3. Eligible list published on 18.06.2025
  4. Demonstration/ document evaluation of Vaccum Extractor conducted on 13.08.2025

I. The reports of technical document evaluation/ demonstration of Vaccum Extractor as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl.No	Equipment	Selected Bidders	Selected Model
1.	Vaccum Extractor	M/s. MEDISIL ENGINEERS	MVDSO2/ YES

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

### 1. M/s. C CUBE ADVANCED TECHNOLOGIES

Rejected model: VAC ASSIST/C CUBE

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
1	Should have Vacuum generator, antibacterial plastic cups, silastic cup and metal cup of different sizes with necessary electrical connection cable	Metal cup not standard
6	Should have traction force indicating display in kilograms	No traction force indicating display
	Comments	Excess noise

### 2. M/s. Kerala Surgical Equipment Co

Rejected model: SVDS 02/Sidra Med

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
6	Should have traction force indicating display in kilograms	Not a dedicated display, traction force indicating display

### 3. M/s. Inscon Medical Soutlions Pvt Ltd

Rejected Model: Barbyl SLD-Bx06/Ceumed Medical

Reason for rejection:



Clause No	Technical Specification	Deviations Noted
3	Should generate a steady vacuum of at least 600mmHg	upto 590 mmHg
6	Should have traction force indicating display in kilograms	Cable connected traction force indicating display

IV. The online price bid of the eligible tenderers mentioned above will be opened on <sup>01.11.2025</sup>~~10.2025~~ @ 11.AM at the etender portal.

Dr Arun P V  
General Manager