



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/7790/2025

Dated: 19.05.2026

NOTICE

Sub:- Tender No.T573/ 2025 - Results of demonstration/ document evaluation & Date of price bid opening – Vein Detecting Transilluminator Device- reg.

- Ref:-**
1. Tender Inviting Notice published dated 14.05.2025
 2. Technical bid opening held on 02.06.2025 , 19.06.2025 ,23.06.2025@ 11.00 A M
 3. Eligible list published on 27.08.2025
 4. Demonstration/ document evaluation of Vein Detecting Transilluminator Device conducted on 12.01.2026

I. The reports of technical document evaluation/ demonstration of Vein Detecting Transilluminator Device as per reference no.4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Selected Models
1	Vein Detecting Transilluminator Device	M/s MEDTRA INNOVATIVE TECHNOLOGIES	VEINEUX AR 100/MEDTRA
		M/s LAKSHMI AGENCY	QV-600/Qualmedi

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. M/s BIOCARE DIAGNOSTICS

Rejected model: NAVI - 60/ MEDCAPTAIN MEDICAL

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Should be able to view veins at 8mm depth	Poor clarity *
6	Should be ideal for adults and infants.	Difficult to detect vein

08.06.2026

III. The online price bid of the eligible tenderer mentioned above will be opened on 09.06.2026 @11.AM at the e-tender portal

Dr Arun P V