KERALA MEDICAL SERVICES CORPORATION LTD
(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600
E-mail id :epetenders.kmscl@kerala.gov.in
CIN : U24233KL2007SGC021616
GSTIN: 32AADCK4029M1ZK

Eqipment Division www.kmscl.kerala.gov.in

No. I/419654/2026 Date:30-10-2026

Notice

Sub:- Tender No.T589/ 2026 - Results of demonstration/ document evaluation & Date of pricebid opening – Iron Cot SS - reg.

Ref:- 1. Tender Inviting Notice published dated 04.03.2026

2. Technical bid opening held on 26.05.2026 @ 11.00 A M

3. Eligible list for demonstration published dated 26.09.2026

4.Demonstration/ document evaluation of Iron Cot SS is conducted on 16.09.2026

I. The reports of technical document evaluation/ demonstration of Iron Cot SS as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1	Iron Cot SS	M/s. Trichur Surgicals	TS 527/Trichur Surgicals
		M/s. GANGOTHRI INDUSTRIES	364B1/GANGOTHRI INDUSTRIES

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A.M/s. MOLECULE WATER TECHNOLOGY

Rejected model; MWT/IC/016/MOLECULE WATER TECHNOLOGY

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	The bed frame should be provided width wise 3 nos of SS 304 angle stiffener of size minimum 35x35x2mm and 2 nos angle support size 50x35x2mm one at each head and leg.	No
6	The leg frames of the cot shall be made of SS 304 tube of diameter 38 mm thick with horizontal bottom tube of diameter 25x1.2 thickness and	No

	vertical filling SS 304 tube of diameter 16mm x 1.2mm thickness.	
9	Angular bracket of 3mm thickness shall be used for fixing the bed frame	No
10	Sockets should be provided for Mosquito net poles 20mmdiameter and 26mm length	No

B. M/s. Samar Steel Industries

Rejected model: SSI-HF 105/Samar Steel Industries

Reason for rejection: Sample not provided

C. M/s. Kerala Surgical Equipment Co

Rejected Model; Medik Series/Kerala Surgical

Reason for rejection: Sample not provided

D. M/s. MEDIMEK INDUSTRIES

Rejected Model: Mi-7013 F/MEDIMEK

Reason for rejection: Sample not provided

E. M/s. Parkins Enterprises

Rejected model; IC 018/Nilkamal Limited

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
4	The top sheet shall be 20 gauge(1mm) thickness SS 304 sheet double bend of height minimum reinforced with channels of size minimum 75mm x 30mm flange welded inside the lengthwise frame bends.	No
5	The bed frame should be provided width wise 3 nos of SS 304 angle stiffener of size minimum 35x35x2mm and 2 nos angle support size 50x35x2mm one at each head and leg.	Not welded
9	Angular bracket of 3mm thickness shall be used for fixing the bed frame	No
10	Sockets should be provided for Mosquito net poles 20mmdiameter and 26mm length	No
12	Should have provisions to connect IV poles on all four corners of the cot.	No
13	Should be provided with leg shoes made of PVC bush of height at least 40mm, outside diameter should be 40mm and thickness at least 4 mm.	No

III. The online price bid of the eligible tenderer mentioned above will be opened on 21.10.2026 @ 11.AM at the e-tender

Rakesh R U

GENERAL MANAGER