

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/8653/2025

Dated: 02.12.2025

NOTICE

Sub:- Tender No.T575(R)/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening –OT Light LED Dual Dome without Camera- reg.

Ref:- 1. Tender Inviting Notice published dated 19.03.2025
2. Technical bid opening held on 13.06.2025 @ 11.00 A M
3. Eligible list for demonstration published dated 21.08.2025
4. Demonstation/ document evaluation of OT Light LED Dual Dome without Camera is conducted on 17.09.2025

I. The reports of technical document evaluation/ demonstration of OT Light LED Dual Dome without Camera as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1.	OT Light LED Dual Dome without Camera	M/s. Lakshmi Surgicals	Karl Storz/ Karl Storz
		M/s. HOSPEDIA MEDICARE PVT. LTD.	LED 50/HOSPEDIA
		M/s. Innovative Concepts	MarLED E9i/E9i/KLS Martin Group
		M/s. Sree Gokulam Healthcare Pvt. Ltd	HyLED C5/C5/Nanjing Mindary

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. M/s. ANTS LIFECARE PRIVATE LIMITED

Rejected model: iLUX 10-10/Xcellance

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
	Comments	Shadow level not acceptable limit

2. M/s. STAAN BIO-MED ENGINEERING PRIVATE LIMITED

Rejected model: DISCOVER/STAAN BIO- MED

Reasonf or rejection:

Clause No	Technical Specification	Deviations Noted
3.b	Light Field diameter shall be above 20 cm or better	Not meeting the specification

3. M/s. BIOCARE DIAGNOSTICS

Rejected model; ECO PLUS 140+140/ENERTECH ELECTRONIC

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3.d	Colour rendering index should not be less than 95	Not meeting the specification

4. M/s. Kerala Surgical Equipment Co

Rejected model: Celesta Series/Eclipse Prism

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3.d	Colour rendering index should not be less than 95	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 08.12 2025 @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager