

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/216/2025

Dated: 11.12.2025

**NOTICE**

**Sub.-** Tender No.T583/ 2025 - Results of demonstration/ technical evaluation & Date of Price Bid Opening –  
Vertical Plasma Freezer -40 Degree C (Deep Freezer -40 Degree) - reg.

**Ref:-** 1. Tender Inviting Notice published dated 07.01.2025  
2. Technical bid opening held on 31.01.2025 @ 11.00 A M  
3. Eligible list published on 18.06.2025  
4. Demonstration/ document evaluation of Vertical Plasma Freezer -40 Degree C (Deep Freezer -40 Degree) conducted on 18.09.2025 & 25.09.2025

I. The reports of technical document evaluation/ demonstration of Vertical Plasma Freezer -40 Degree C (Deep Freezer -40 Degree) as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model/OEM
1.	Vertical Plasma Freezer -40 Degree C (Deep Freezer -40 Degree)	M/s. KOCHIN INLAB EQUIPMENTS INDIA PVT LTD	DF40U/TERUMO PENPOL

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. M/s. Labline Instruments  
Rejected model:VDF 2853/ LABLINE  
Reason for rejection: Sample not produced
2. M/s. Leads India Laboratory Equipment Co  
Rejected model: LPF-350/LEADS INDIA (LABTOP)  
Reason for rejection; Sample not produced
3. M/s. Rattan Sales Corporation  
Rejected model: 1180-SFR-042/ Rattan Sales  
Reason for rejection: Sample not produced
4. M/s. Instromedix India Pvt. Ltd  
Rejected model: KBS D 350/ Knish Biomedicals  
Reason for rejection: Sample not produced
5. M/s. New India Surgicals and Chemicals  
Rejected model: PF-V-325/CryoScientific  
Reason for rejection;

Clause No	Technical Specification	Deviations Noted
19	Door Open: Audio-Visual	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 16/12/2025 @ 11.00 am at the e-tender portal.

Dr Arun P V  
General Manager