



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/15172/2024

Dated: 16.12.2024

NOTICE

Sub:- Tender No.T548/ 2024 - Results of demonstration/ document evaluation & Date of price bid opening – Digital Hemoglobinometer - reg.

Ref:- 1. Tender Inviting Notice published dated 27.08.2024
2. Technical bid opening held on 23.09.2024 @ 11.00 A M
3. Demonstration/ document evaluation of Digital Hemoglobinometer conducted on 26.11.2024, 04.12.2024 & 12.12.2024

I. The reports of technical document evaluation/ demonstration of Digital Hemoglobinometer as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

| Sl. No | Item | Name of the Selected Bidder | Model/OEM |
|--------|--------------------------|-----------------------------|-------------------------------------------------|
| 1. | Digital Hemoglobinometer | M/s. Wrig nanosystem | TrueHB/Wrig Nanosystems |
| | | M/s. Thirupathi Enterprises | DOLPHIN HEMOGLOBIN METER/ADITYA CLINICAL SYSTEM |

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s. Peerless Biotech Pvt. Ltd

Rejected model: Quick Check Plus/ Acon Biotech

Reason for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|----------------------------------------------------------------|------------------|
| 3 | Internally rechargeable battery /External rechargeable battery | Not available |
| III.d | specificity: 100% | 86.70% |

B. M/s. Biosense Technologies Private Limited

Rejected model; HbChek V2.0/Biosense technologies Private Ltd

Reason for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|-------------------------|------------------|
| III.d | specificity: 100% | 80% |

C. M/s. MICROGENE DIAGNOSTIC SYSTEMS PVT. LTD

Rejected model: AccuSure HB/Macrad Medical Private Limited

Reason for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|-------------------------|------------------|
| III.d | specificity: 100% | 86.70% |

D. M/s. Sensa Core Medical Instrumentation Pvt Ltd

Rejected model; Hemo Spark / Sensacore

Reason for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|-------------------------|------------------|
| III.d | specificity: 100% | 86.70% |

E. M/s. DEEP MEDITECH PRIVATE LIMITED

Rejected model: Aspen HB/ Aspen

Reason for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|-------------------------|------------------|
| III.d | specificity: 100% | 60% |

III. The online price bid of the eligible tenderer mentioned above will be opened or ^{18.} 12.2024 @11.00 am at the e-tender portal.

Dr. Shibulal A
General Manager