



#### KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600

E-mail id: <a href="mailto:eptenders.kmscl@kerala.gov.in">eptenders.kmscl@kerala.gov.in</a>
CIN: U24233KL2007SGC021616
GSTIN: 32AADCK4029M1ZK

**EP Division** 

www.kmscl.kerala.gov.in

No: KMSCL/EP/10408/2024

Dated: 05.04.2025

#### NOTICE

Sub:- Tender No.T544(R)/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment- reg.

Ref:- 1. Tender Inviting Notice published dated 11.06.2024

- 2. Technical bid opening held on 15.07.2024 @ 11.00 A M
- 3. Demonstation/ document evaluation of various equipment conducted on 04.02.2025 & 06.03.2025

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI. No	Equipment	Name of the Selected Bidder	Selected ModeL
		M/s. Kerala Surgical Equipment Co.	PSI Levitate HD/Palakkad
1.	Surgeons Chair	M/s. C CUBE ADVANCED TECHNOLOGIES	C-143C/C CUBE
2		M/s. Instromedix India Pvt. Ltd	X12 and MFM -CMS/Edan Instruments
2.	Monitoring System	M/s.Sree Gokulam Healthcare Pvt.	uMec7/Shenzhen Mindray
3.	Hand Held Pulse Oximeter	M/s. SUMMIT HEALTHCARE PVT.	BIOLIGHT -
		LTD.	M800/GUANGDONG
		M/s. Sree Gokulam Healthcare Pvt. Ltd	PM60/Shenzen Mindray
		M/s. Biomedical Engineering Company	RAD G/Masimo Corporation
		M/s. BIOCARE DIAGNOSTICS	MD300K2/Beijing Choice Electronic

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

### A.KMSCL/EP/T544(R)/1697/2024: Surgeons Chair

1. M/s. Dynamic International

Rejected model; DHF-2361/DYNAMIC

Reason for rejection: Sample not produced

2. M/s. STAAN BIO-MED ENGINEERING PRIVATE LIMITED

Rejected model; Throne/STAAN

Reason for rejection: Sample not produced.

# B. KMSCL/EP/T544(R)1087/2024: Multiparameter Monitor (5 Para) With Central Monitoring System

1. M/s.Allied Medical Ltd

Rejected model: Meditec M777/Allied Medical Limited

Reason for rejection: Sample not produced

2. M/s. Clarity Medical Pvt Ltd

Rejected model:Spectra Gold/Clarity Medical

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
II	Central Monitoring System	Not demonstrated

### 3. M/s. SUMMIT HEALTHCARE PVT. LTD

Rejected model: BIOLIGHT - Anyview S12/GUANGDONG BIOLIGHT

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
11	Central Monitoring System	Not demonstrated

#### 4. M/s.Blaze Systems and Services

Rejected model: C80/Shenzhen Comen Medical

Reason for rejection:

Clause No	Technical Specification Deviations Noted	
	Should have safety certificate from a	
	competent authority CE issued by a notified	
18	body registered in the European commission No valid certificate	
	/ FDA (US). The certificate attached shall be	
	verifiable.	

## 5. M/s.UNIVERSAL AGENCIES

Rejected model:K15/Shenzhen Creative

Reason for rejection: Sample not produced

### C. KMSCL/EP/T544(R)/105C/2024: Hand Held Pulse Oximeter

### 1. M/s. UNIVERSAL AGENCIES

Rejected model: UNI-EM 810A/ELECTROMEDICALS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
16	Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced	No valid certificate
	along with the technical bid	

# 2. M/s. S.S. Medical Systems India Pvt. Ltd

Rejected model: SONO-OX-PL/S S Medical Systems

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Should have safety certificate from a	
	competent authority CE issued by a	
	notified body registered in European	
	Commission / FDA (US) / STQC CB	
16	certificate / STQC S certificate or valid	No valid certificate
	detailed electrical and functional safety	
	test report from ERTL. Copy of the	
	certificate / test report shall be produced	
	along with the technical bid	

III. The online price bid of the eligible tenderer mentioned above will be opened on  $\mathfrak{S}$ 0.04.2025 @ 11.AM at the e-tender portal.

**Dr. Shibulal A**General Manager

