



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10401/2024

Dated: 24.06.2025

NOTICE

Sub:- Tender No.T541/ 2024 -Cancellation Notice – Fume Hood - reg.

Ref:- 1. Tender Inviting Notice published dated 11.06.2024

2. Technical bid opening held on 29.11.2024 @ 11.00 A M

3. Reports of demonstration/ document evaluation of Fume Hood received on 30.04.2025

I. The reports of technical document evaluation/ demonstration of Fume Hood as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Fume Hood	M/s LABLINE INSTRUMENTS	FC-3/LABLINE

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s. UNIQUE SCIENIFIC

Rejected model: LOL MASTER700/CARL Zeiss

Reason for rejection:

Clause No	Technical Specification	Deviations noted
4.	The working table should be covered with acid proof glazed tiles, provided with FRP moulded / ceramic sink (acid corrosion resistant material) and water tap, provided with a cock for gas or air, additional power point, diffused florescent tube.	Product catalogue mentions work table is of granite but in amendment granite slab is not considered.
9	Approximate external dimensions in cm 96 (W) x 75 (D) x 180 (H)	96x75x240

B. M/s. Kerala Surgical Equipment

Rejected model: PFH-32/Rotek plus

Reason for rejection:

Clause	Technical Specification	Deviations noted
--------	-------------------------	------------------

No		
9	Approximate external dimensions in cm 96 (W) x 75 (D) x 180 (H)	not specified

C M/s. Instrumedix India Pvt Ltd

Rejected model; PLT-272(H)/P.L Tandon & company

Reasons for rejection:

Clause No	Technical Specification	Deviations noted
4.	The working table should be covered with acid proof glazed tiles, provided with FRP moulded / ceramic sink (acid corrosion resistant material) and water tap, provided with a cock for gas or air, additional power point, diffused florescent tube.	Product catalogue mentions work table is of granite but in amendment granite slab is not considered.

D. M/s. OCEAN LIFE SCIENCE COPERATION

Rejected model; OLSC-111-3/OCEAN LIFE

Reason for rejection:

Clause No	Technical Specification	Deviations noted
9	Approximate external dimensions in cm 96 (W) x 75 (D) x 180 (H)	Product catalogue gives inner dimension as 96x75x180. External dimension not specified.

E. M/s UNIVERSAL AGENCIES

Rejected model; YSI-225/YOURK SCIENTIFIC

Reason for rejection:

Clause No	Technical Specification	Deviations noted
9	Approximate external dimensions in cm 96 (W) x 75 (D) x 180 (H)	96x75x240.

III. Due to administrative reasons the tender for **KMSCL/EP/T541/785A/2024:Fume Hood** is cancelled.

Dr Arun P V
General Manager