

Running Contract Details	
Equipment Name	Power drill and reamer for Orthopaedics
Running Contract Valid Till	14-11-2026
Tender Ref No	KMSCL/EP/T544/104/2024
Tendered Quantity	50
Supplier Name	M/s S H Pitkar Orthotools Pvt Ltd
GST No	27AACCS6013Q1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
EL 32 J Block MIDC Bhosari Pune-411026	Contact Person	
	Phone	
	Mobile No	
	Email	info@pitkar.net,pitkar.marketing3@gmail.com,pitkar.praseetha@gmail.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Power drill and reamer for Orthopaedics <i>Model & Make : pitkar/S.H. Pitkar</i>			231702.02 Incl.GST :6%	17088.02	248790.04	
				231702.02	17088.02	248790.04	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Power drill and reamer for Orthopaedics							
Labour	4,137.54	4,137.54	4,137.54	4,137.54	4,137.54	4,137.54	4,137.54
Comprehensive	8,275.07	10,343.84	12,412.61	14,481.38	16,550.15	18,618.91	20,687.68

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Power drill and reamer for Orthopaedics

Technical Specification

Equipment: Power Drill and Reamer for Orthopedics

1. Should be supplied with a driving unit including motor, stand, foot control, flexible shaft, tool kit and container.
2. Should have a cannulated drill hand piece of maximum speed 1200rpm with Non-corrosive Jacob chuck (0-1/4"). Autoclavable.
3. Should have a cannulated reaming hand piece with maximum speed 400rpm with AO type quick coupling. Autoclavable.
4. Should have a pistol grip sagital saw with a set of 5 blades. Autoclavable.
5. Should have a flexible reamer shaft 8mm diameter fixed head.
6. Should have a flexible reamer shaft for detachable heads up to 12mm.
7. Should have reamer heads from 8mm to 12 mm (set of 9).
8. Should have reamer heads from 12.5mm to 15mm (set of 6).
9. Should have flexible reamer shaft for detachable heads up to 15mm. Autoclavable.
10. Should have an extra flexible shaft for driving unit autoclavable.
11. Should work with input 200 to 240Vac 50 Hz supply.