



Running Contract Details	
Equipment Name	Infantometer
Running Contract Valid Till	08-04-2027
Tender Ref No	KMSCL/EP/T544/592/2024
Tendered Quantity	50
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Infantometer <i>Model & Make : PSI Levitate HD/Palakkad</i>	2336.4 Incl.GST :18%	163.55	2499.95
		2336.4	163.55	2499.95

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Infantometer

Technical Specification

Equipment: Infantometer

1. Portable baby/infant length-height measuring system
2. Measures laying length of neonates and babies
3. No need for calibration as all parts have prefixed position
4. Reads in centimeters and inches
5. Minimum graduation: 1 mm
6. Long-lasting hard-wearing ruler/graduation is fully integrated with device
7. Measuring slide/wedge glides smoothly and close via ruler, avoiding reading parallax
8. Measuring slide/wedge wobbles max 2 mm, over full length
9. No sharp edges or corners
10. Low stable board, width: ca 30 cm
11. Length, measurement range, approx: 100 cm
12. Head/footplate, board and slide/wedge made of quality laminated wood or plastic
13. Wood parts should be treated and finished/protected with varnish to prevent chipping of edges and allow easy cleaning; all connections should be screwed/nailed plus glued