



Running Contract Details	
Equipment Name	Dry Bath
Running Contract Valid Till	27-06-2026
Tender Ref No	KMSCL/EP/T521/315/2024
Tendered Quantity	25
Supplier Name	M/s AASTMEDI TECH PRIVATE LIMITED
GST No	32AATCA5385Q1ZT
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Door No 16/365 16/366 16/367 Sreevihar Pallarimangalam Olakkettyambalam P.O. Mavelikara	Contact Person	Prasanth G Pillai
	Phone	8589019182
	Mobile No	9567859180
	Email	info@astmeditech.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Dry Bath <i>Model & Make : DDI 24/AAST MEDITECH</i>	26432 Incl.GST :18%	1850.24	28282.24
		26432	1850.24	28282.24

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Dry Bath

Equipment: Dry Bath

1. Temperature range: ambient to 110 degree C. Controlled to ± 0.2 deg. C (or better) by Solid State Time Proportional Temperature Controller.
2. Digital-Display using PT-100 Sensor
3. Should be supplied with 3 Nos. interchangeable and autoclavable anodized aluminium blocks of 75x50x50 mm size for accommodating 0.5 ml, 1.5 ml tubes
4. 12mm, 16mm or 19mm test tubes or two block of 150x75x50 mm size for accommodating 100 Tubes of 10mm Diameter
5. The body made of heavy M.S. and painted with epoxy powder paint.
6. All accessories needed to make the equipment functional, must be supplied