



Running Contract Details	
Equipment Name	Cryosurgery Unit
Running Contract Valid Till	28-10-2026
Tender Ref No	KMSCL/EP/T539/26/2024
Tendered Quantity	10
Supplier Name	M/s Inscon Medical Solutions Private Limited
GST No	32AACCI7233A1Z8
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
98C Manthrakkal Junction Thaikkattukara Post Aluva Kochi- 683106	Contact Person	ajesh
	Phone	
	Mobile No	9744160009
	Email	sales@inscon.in

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Cryosurgery Unit <i>Model & Make : Wellach LL 100/ Cooper Surgicals</i>			420000 Incl.GST :12%	30975	450975	
				420000	30975	450975	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Cryosurgery Unit							
Labour	12,000.00	12,000.00	12,000.00	13,000.00	14,000.00	15,000.00	16,000.00
Comprehen sive	25,000.00	25,000.00	30,000.00	30,000.00	35,000.00	40,000.00	50,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cryosurgery Unit

Technical Specification

Equipment: Cryosurgery Unit for Gynecology

1. Should be stand-alone unit consist of a tank, a pressure regulator, and a probe attached by tubing to the tank.
2. Should be a fully mobile cryo-surgical system with a wide array of interchangeable probes designed for the use of gynaecologic surgical procedures. The interchangeable probes must include both different sizes for the cervix besides flat one for vaginal lesions.
3. Either Nitrous oxide or carbon dioxide can be used as refrigerant.
4. Units should support various probes and tips.
5. Nitrous based unit should have scavenging ability.
6. Should have adjustable freezing temperatures and pressures through a regulation system.
7. Should have non-electric defrosting system.
8. All cryo probes must be autoclavable.
9. Operating pressure 450 to 800 psi.
10. Operating temperature -70 to -10°C for Carbon dioxide and -90 to -25°C for Nitrous oxide.
11. Should be supplied with triggers/connection for N2O or CO2.
12. Should be supplied with rolling cart.
13. Should be supplied with unfilled cylinder for N2O or CO2.
14. Should be supplied with all kinds of probes required for gynecology