



Running Contract Details	
Equipment Name	Endo Laser Unit
Running Contract Valid Till	20-11-2026
Tender Ref No	KMSCL/EP/T544/1701A/2024
Tendered Quantity	10
Supplier Name	M/s Alcon Laboratories (India) Private Limited
GST No	29AACCA3430F1Z2
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
11th Floor RMZ Azure Bellary Road Hebbal Bengaluru - 560092 Karnataka India	Contact Person	V. Nandagopal
	Phone	080-40064600 , 4544 4444
	Mobile No	+91 9790967992
	Email	alcon.tenders@alcon.com,nandagopal.v@alcon.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Endo Laser Unit <i>Model & Make : purepoint Laser/Alcon Laboratories</i>			2718100 Incl.GST :12%	200459.88	2918559.88	
				2718100	200459.88	2918559.88	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Endo Laser Unit							
Labour	1,11,062.00	1,11,062.00	1,11,062.00	1,11,062.00	1,11,062.00	1,11,062.00	1,11,062.00
Comprehensive	1,85,485.00	1,85,485.00	1,85,485.00	1,85,485.00	1,85,485.00	1,85,485.00	1,85,485.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Endo Laser Unit

Technical Specification

Equipment:

1. The system should have Green Laser capability
2. The Green Laser should have dual port option
3. Should have a multifunction foot pedal to change power and Ready/Standby modes directly by footswitch
4. The system should have voice confirmation
5. The system should have RFID capacity which recognizes the probe connected and automatically load the settings
6. Endolaser probe - flexible with adjustable articulation and compatible for 23 and 25 gauge vitrectomy