



Running Contract Details	
Equipment Name	Gynec OT Table
Running Contract Valid Till	28-05-2027
Tender Ref No	KMSCL/EP/T544/293L/2024
Tendered Quantity	30
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	<b>Gynec OT Table</b> <i>Model &amp; Make : Medik Comfort 60/Medik Enterprises</i>			247800 Incl.GST :18%	17346	265146	
				<b>247800</b>	<b>17346</b>	<b>265146</b>	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
Gynec OT Table							
Labour	4,200.00	4,200.00	4,200.00	4,200.00	4,200.00	4,200.00	4,200.00
Comprehensive	14,700.00	14,700.00	14,700.00	14,700.00	14,700.00	14,700.00	14,700.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Gynec OT Table**

Technical Specification

#### **Equipment: Gynec OT Table**

1. The table should be electrically operated.
2. It should have overall approximate size 1830 (l) x 560 (w)mm without accessories
3. The table top should have a minimum height of 680mm , vertical stroke of minimum of 300mm
4. Should have at least 25° trendelenburg and reverse trendelenburg
5. Should have at least 15° lateral tilt movements
6. Should have at back up movement with gas spring mechanism of 65° up/ 25° down.
7. Should have head rest, hand rest and leg rest detachable with fixing clamps.
8. Should have SS/ABS side rails with safety guard for clamps on both sides of the table, column covers the electro mechanical actuator system
9. Table top should be fitted with X-ray permeable top
10. The table should have heavy metal base covered with molded ABS/SS.
11. The table should be fitted with non rusting 50mm dia casters, foot lever operated table lock
12. All MS parts used in the table should be pretreated and epoxy powder coated
13. Should be supplied with one pair of leg crutches with side rail locks, shoulder rest (one pair) with side rail locks, side support with lock.
14. Cushions' mountable and 3 pieces- head, body and foot with minimal gap