

Running Contract Details	
Equipment Name	Dental X-ray machine
Running Contract Valid Till	28-10-2026
Tender Ref No	KMSCL/EP/T529/31/2024
Tendered Quantity	30
Supplier Name	M/s Iatome Electric India Pvt Ltd
GST No	33AACCI3294D1ZV
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
S.F.No.61/1B 61/2 Aadhavan Industrial Estate Athipalayam Road Chinnavedampatty Coimbatore -641049	Contact Person	Ms.Saravanakumar
	Phone	1800-833-8090
	Mobile No	9943465551
	Email	sales@alerio.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Dental X-ray machine <i>Model & Make : Alerio DC Optima/ Iatome Electric India Pvt Ltd</i>	150000 Incl.GST :12%	11062.5	161062.5
		150000	11062.5	161062.5

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Dental X-ray machine							
Labour	6,000.00	7,000.00	8,000.00	9,000.00	10,000.00	11,000.00	12,000.00
Comprehensive	15,000.00	17,000.00	19,000.00	21,000.00	23,000.00	25,000.00	27,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Dental X-ray machine

Technical Specification

Equipment: Dental X- Ray Machine

1. Generator Constant Potential (DC)
2. Focal Spot size 0.8 mm or better
3. Tube voltage Selectable 65 KVP fixed or better
4. Current 7mA fixed or better
5. Exposure time 0.01sec to 1 sec or better
6. Arm Length At least 80"
7. Line Voltage 220-230Vac
8. Accessories Round 8" cone Stand Model with four legs and Wheels
9. Should Supply a. Lead Apron with Thyroid guard – Light weight.
10. The offered model of X ray machine should have AERB type approved. Copy of the certificate should be submitted along with the offer form.