



Running Contract Details	
Equipment Name	MR safe portable multi-parameter vital sign monitor
Running Contract Valid Till	15-02-2026
Tender Ref No	KMSCL/EP/T501(R)/1666/2023
Tendered Quantity	6
Supplier Name	M/s Biocare Diagnostics
GST No	32AAOFB5438P1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	5 Years

Supplier`s Details		
Address	Contact Details	
787a-1 First Floor Kamala Building Chalappuram (Po) Calicut 673002	Contact Person	Linesh.T
	Phone	
	Mobile No	9847009640
	Email	mail@biocarecalicut.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	MR safe portable multi-parameter vital sign monitor Multifunctional couch with IV fluid stand as per specification" <i>Model & Make : 3880 with other modules & accessories / IRADIMED USA</i>	8749440 Incl.GST :12%	645271.2	9394711.2
2	SPO2 probe (Pediatric& Adult) Rate for future purchase	50400 Incl.GST :12%	3717	54117
3	BP Cuff (Pediatric& Adult) Rate for future purchase	22400 Incl.GST :12%	1652	24052
		8822240	650640.2	9472880.2

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year	11 th Year	12 th Year
MR safe portable multi-parameter vital sign monitor Multifunctional couch with IV fluid stand as per specification"							
Labour	4,00,000.00	4,00,000.00	4,00,000.00	4,00,000.00	4,00,000.00	4,00,000.00	4,00,000.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Comprehensive	9,50,000.00	9,75,000.00	10,00,000.00	10,25,000.00	10,50,000.00	11,00,000.00	11,25,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :MR safe portable multi-parameter vital sign monitor Multifunctional couch with IV fluid stand as per specification"

Sl. No.	Item Description
1.	It should be a fully Non-Magnetic multi parameter portable patient monitoring solution, designed to be small, easy to use and lightweight. monitor should have at least 10 inch display.
2.	MRI vital signs monitor able to travel with the patient
3.	It should be 30000 Gauss complaint so that it can be placed anywhere in the MR room (up to 3 Tesla MRI)
4.	The unit should come with wireless vital signs 3 or 5 lead ECG with trusted artifact free spo2 technology.
5.	The unit should come with wireless control room light weight monitor with base station having back up charging dock.
6.	Clinical Features: Standard
7.	SpO2 with perfusion Indicator: Wireless
8.	ECG:3/5 Lead: Wireless
9.	Non- Invasive Blood Pressure
10.	Dynamic Trend indicators
11.	Tri-Colored alarm light
12.	Full gas module with ETCO2 Side stream. Etco2 sample lines 100 plus 50 water trap/water locks to be supplied along with the machine
13.	Invasive Blood Pressure
14.	Accessories, MR Compatible Laryngoscope. MR compatible miller blades 0,1 and Macintosh blades 0,1 2,3,4
15.	Vendor should provide the Pediatric& Adult SPO2 probes-03 each
16.	Vendor should provide the Pediatric& Adult BP Cuff-03 each
17.	Vendor should also quote the price for SPO2 probe and BP Cuff (Pediatric& Adult) separately for the further purchase if required.
18.	Should be European CE marked/FDA marked

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall

prevail