

Running Contract Details

Equipment Name	Patient warmer
Running Contract Valid Till	31-12-2026
Tender Ref No	KMSCL/EP/T519(R)/165/2023
Tendered Quantity	100
Supplier Name	M/s STANDARD MEDICAL CORPORATION
GST No	29AEIFS1680L1ZF
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details

Address	Contact Details	
No. 6 Madake Swamy Mutt Road Kengeri Bengaluru -560060	Contact Person	Mr Sinsha TS
	Phone	
	Mobile No	9633066542
	Email	info@stanmed.in

Item-wise Price Details

#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Patient warmer <i>Model & Make : IOB-505/IOB Medical Inc, USA</i>	75600 Incl.GST :5%	5947.2	81547.2
2	Cost of additional Blanket disposable adult full body (for future purchase)	458.85 Incl.GST :5%	36.1	494.95
3	Cost of additional Blanket disposable adult Upper half (For future purchase)	458.85 Incl.GST :5%	36.1	494.95
		76517.7	6019.39	82537.09

Annual / Comprehensive Maintenance Charges (Exl.Tax)

Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Patient warmer							
Labour	2,160.00	2,268.00	2,381.40	2,500.47	2,625.49	2,756.76	2,894.00
Comprehensive	4,320.00	4,536.00	4,762.80	5,000.94	5,250.99	5,513.54	5,789.22

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Patient warmer

1. Microprocessor Based.
2. Unique Temperature Control System.
3. High Air Flow Temperature range: 34- 42°C
4. Alarm indicator for Over Temperature and Under Temperature.
5. Temperature Digital Display.
6. In-built Air Filter.
7. Elapsed Time Display.
8. IV pole/bedrail is mountable.
9. Accessories :
 - a. 3 Nos. of blanket disposable adult full body
 - b. 3 Nos. of the blanket disposable upper half
10. Should work on 220-240V AC/50Hz.
11. Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission/FDA(US)/STQC CB certificate /STQC S certificate or a valid detailed electrical and functional safety test report from ERTL. Copy of the certificate /test report shall be produced along with the technical bid

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail