

Running Contract Details	
Equipment Name	Surgeons Chair Motorized
Running Contract Valid Till	24-10-2026
Tender Ref No	KMSCL/EP/T451/88B/2022
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Surgeons Chair Motorized</b> <i>Model &amp; Make : MEDIK- SCM-85 / Medik Enterprises</i>	69620 Incl.GST :18%	4873.4	74493.4
		<b>69620</b>	<b>4873.4</b>	<b>74493.4</b>

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Surgeons Chair Motorized**

1. Should have motorized up and down movement
2. Should have foot operated control.
3. Should have an adjustable back support (Front & Back Movement).
4. Should have height adjustable / detachable arm rests.
5. Should be made up of SS304 Grade or Powder coated mild steel completely with upholster seat, back rest and arm rest
6. Minimum height should be 20" and maximum should be 28".
7. Should have full swiveling castors for easy mobility. Castors with brake / Braking mechanism.