

Running Contract Details	
Equipment Name	Plasma Thawing bath
Running Contract Valid Till	31-03-2025
Tender Ref No	KMSCL/EP/T450/164/2022
Tendered Quantity	15
Supplier Name	M/s New India Surgical and Chemicals
GST No	32AAGFN8445E1ZD
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
30/1212IV Prasanth Buildings Near Medical College Hospital Calicut-673008	Contact Person	Prasanth N
	Phone	0495-2354188
	Mobile No	9447054188
	Email	nisc2000@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Plasma Thawing bath <i>Model & Make : SAFE thaw/ Cryo Scientific</i>	106082 Incl.GST :18%	7425.74	113507.74
		106082	7425.74	113507.74
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Plasma Thawing bath				
Labour	5,000.00	5,000.00	5,000.00	
Comprehensive	15,000.00	15,000.00	15,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Plasma Thawing bath

1. Should be able to thaw a minimum of 12 plasma bags (FFP bags)
2. Should be a water bath based system operating at a preset and precise temperature of 37°C.
3. Should have separate slot for holding the regular plasma bags.
4. Should have audio /visual over-temperature alarm system
5. Should have a deep thawing chamber with a stirrer/ pump
6. Should have a system to drain the chamber.
7. Should be supplied with a cover to keep the unit covered when not in use
8. Should have display for temperature.
9. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid