



Running Contract Details	
Equipment Name	AED Trainer with Simulator
Running Contract Valid Till	26-01-2028
Tender Ref No	KMSCL/EP/T573/1782/2025
Tendered Quantity	100
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
43/2675A Dhanya Complex Sastha Temple Road Ernakulam North PO Kochi- 682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	AED Trainer with Simulator <i>Model & Make : BeneHeart C2/ Shenzhen Mindray</i>	30450 Incl.GST :5%	2395.4	32845.4
		30450	2395.4	32845.4

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :AED Trainer with Simulator

1. General Requirements:

- a. The AED Trainer must be designed for realistic training in the use of an Automated External Defibrillator (AED) without delivering actual shocks.
- b. The device should replicate the most common features of AEDs available in the market to ensure practical hands-on training.
- c. It should be suitable for training healthcare professionals, first responders, and laypersons in emergency scenarios.
- d. The trainer must be compliant with international guidelines such as AHA, ERC, and ILCOR.

2. Functional Requirements:

- a. Must provide audio and visual prompts for step-by-step guidance during CPR and AED usage.
- b. The trainer should simulate different emergency scenarios, including shockable and non-shockable rhythms.
- c. Should include a metronome feature for guiding correct compression rates during CPR.
- d. The device must have pre-configured training scenarios that can be controlled via remote control or manual selection.
- e. Should allow for hands-on pad placement training with reusable adhesive training pads.
- f. Capability to train on fully automatic and semi-automatic AED modes.

3. Physical and Technical Specifications:

- a. The AED Trainer must have a durable, lightweight, and portable design.
- b. Must include a clear, high-contrast display or LED indicators for ease of use.
- c. Training pads should be compatible with adult and Pediatric mannequins.
- d. Device should operate on rechargeable or replaceable batteries, with a power-efficient design.
- e. Must include a battery status indicator.
- f. Remote control should be included for instructors to simulate different scenarios easily.
- g. Must include multilingual support as per requirement.

4. Safety and Compliance:

- a. The trainer should not deliver any real defibrillation shocks.
- b. All components should be made from non-toxic, latex-free materials.
- c. The device must comply with relevant safety standards such as CE/ FDA/BIS or equivalent.

5. Accessories and Inclusions:

- a. AED Trainer unit Remote control for scenario selection/ App for Scenario Selection
- b. Carry case for portability
- c. User manual and instructional guide
- d. Power adapter or batteries