



Running Contract Details	
Equipment Name	Urine Toxicology Analyser
Running Contract Valid Till	28-10-2026
Tender Ref No	KMSCL/EP/T371/1375B/2020
Tendered Quantity	25
Supplier Name	M/s R.K.Agency
GST No	32AKWPJ3317R1ZJ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
43/713 Cemetery Junction Pothady House Power House Extension Road Ernakulam - 18	Contact Person	Mr. Jitendra Bhandari
	Phone	
	Mobile No	9746015353
	Email	rkagencykerala@rkagency.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Urine Toxicology Analyser <i>Model & Make : Quidel Triage Analyzer/Quidel</i>	354000 Incl.GST :18%	24780	378780
2	Cost of One test card	1120 Incl.GST :0%	92.51	1212.51
		355120	24872.51	379992.51

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Urine Toxicology Analyser							
Labour	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Comprehensive	30,000.00	30,000.00	30,000.00	30,000.00	30,000.00	30,000.00	30,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Urine Toxicology Analyser

Urine Toxicology Analyzer

1. Should be able to measure Qualitative/ quantitative measurement of Urine Toxicology
2. Able to measure a minimum of 10 parameters with a single test card. The parameters should be able to measure - APAP, AMP, mAMP, BAR, BZO, COC, MTD, OPI, PCP, THC, TCA)
3. It should be a portable unit and have an inbuilt battery backup.
4. It should have a parameter-wise QC checking facility.
5. Should have a memory capacity of minimum 200 test results with date time and comments.
6. Should have minimum 200 code chip records (100 test + 100 QC).
7. Should be able to use a minimum sample volume of approximately 250 microlitres.
8. Should work on 200 – 240 Vac 50Hz power supply.
9. It should have an internal printing facility or should be supplied with an external thermal printer.
10. Should be provided with a calibration certificate issued by the manufacturer at the time of installation.

Note: The rate for the test card should be mentioned in the BOQ and this rate shall be taken for evaluation. The rate shall be fixed for 3 years from the date of price bid opening.