

Running Contract Details	
Equipment Name	OT Table Electro Hydraulic
Running Contract Valid Till	03-06-2026
Tender Ref No	KMSCL/EP/T522/197A/2024
Tendered Quantity	100
Supplier Name	M/s STAAN BIO-MED ENGINEERING PRIVATE LIMITED
GST No	33AAICS8026R1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	9,548% & 0 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
190-A BHARATHIAR ROAD GANAPATHY COIMBATORE TAMIL NADU-641006	Contact Person	Mr M MARIA ALBERT STANLY
	Phone	0422 - 2531008
	Mobile No	9363121030
	Email	staan3@hotmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>OT Table Electro Hydraulic</b> <i>Model &amp; Make : GLORY/STAAN</i>	880000.34 Incl.GST :18%	61600.02	941600.36
2	<b>One Horse shoe shaped head rest (ss grade 304)</b>	29500 Incl.GST :18%	2065	31565
3	<b>One adapter to fix three pin skull clamp(ss grade 304)</b>	14160 Incl.GST :18%	991.2	15151.2
4	<b>One Hand Operating table with height adjustable floor support</b>	23600 Incl.GST :18%	1652	25252
5	<b>One raised arm Board with lock</b>	11800 Incl.GST :18%	826	12626
6	<b>One Gel Body Roll for adult(Size-510 mmL X 155mmW X 125mmH)</b>	5900 Incl.GST :18%	413	6313

Item-wise Price Details							
7	One Electrocautery earthing sheet			11800 Incl.GST :18%	826	12626	
				976760.34	68373.22	1045133.56	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
OT Table Electro Hydraulic							
Labour	22,500.00	22,500.00	22,500.00	22,500.00	22,500.00	22,500.00	22,500.00
Comprehensive	37,500.00	37,500.00	37,500.00	37,500.00	37,500.00	37,500.00	37,500.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :OT Table Electro Hydraulic**

#### **Equipment Name: OT Table Electro Hydraulic**

1. The table should have minimum of 4 sections ie. head section, leg section, seat section and back plate section.
2. The table should be electro hydraulically operated having the following hand switch operated electro hydraulic functions (all the dimensions will have a permitted deviation of +/- 10 %)

Sl No	Description	Range
1.	Up / Down	680-1000mm
1.	Trendelenberg & Reverse	30deg
1.	Side Tilting (Lateral)	Should be electro hydraulically operated - 30 Deg
1.	Zero Position (Return to Normal)	
1.	Top Slide	300mm
1.	Braking (by hand switch)	
1.	Back Plate (Sitting Position)	-40 to +80 should be electro

1. All the above functions must have manual override Override System present on the column for up/down, Trendelenburg /reverse Trendelenburg and tilt, Floor locking.
2. In addition to the above hand switch operated functions, the table must have the following manual functions.

	Description	Range
1.	Head Section Tilting	30deg Up / 90deg Down
1.	Split Leg Plate manual movement	+10 to -90

1. The table top must be made of durable radiolucent bakelite material capable of withstanding exposure to frequent C-Arm imaging, without diminishing the image clarity.
2. The table should be supplied with the following accessories.
  - i. Mattress for the complete table top in sections - 1 set
  - ii. A pair of arm boards with pad and fixing clamp - 1 set
  - iii. A pair of padded shoulder support with clamps (SS grade 304) – 1
  - iv. A pair of padded lateral support with clamps (SS grade 304) – 1
  - v. Anesthetic screen frame with clamp (SS grade 304) – 1
  - vi. Patient restraint strap – 1
  - vii. leg crutches with side rail locks – 1 pair
1. The base cover, lifting column cover and side rails should be made of stainless steel grade SS grade 304
2. Should have enhanced weight bearing casters fitted with ball bearing.
3. The table should have a heavy and sturdy base and compact to provide adequate foot room for the operating team.
4. The weight bearing capacity of the table shall be at least 175kg.
5. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
6. Battery backup shall be available for minimum 50 functions.
7. Should have following accessories and the rate for the same should be offered separately.
  - a. Horse shoe shaped head rest (SS grade 304) and adapter to fix three pin skull clamp (SS grade 304) - 1 set
  - b. Hand Operating Table with height adjustable floor support - 1 set
  - c. Raised arm Board with lock - 1 No
  - d. Gel Body Roll for adult - 1 No (Size 510mm L X 155mm W X 125mm H)
  - e. Electrocautery earthing sheet – 1 No.