



Running Contract Details	
Equipment Name	Posterior Segment - Victrectomy system
Running Contract Valid Till	01-12-2026
Tender Ref No	KMSCL/EP/T544/258/2024
Tendered Quantity	5
Supplier Name	M/s Alcon Laboratories (India) Private Limited
GST No	29AACCA3430F1Z2
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
11th Floor RMZ Azure Bellary Road Hebbal Bengaluru - 560092 Karnataka India	Contact Person	V. Nandagopal
	Phone	080-40064600 , 4544 4444
	Mobile No	+91 9790967992
	Email	alcon.tenders@alcon.com,nandagopal.v@alcon.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Posterior Segment - Victrectomy system <i>Model & Make : constellation Tabletop (TT) / Alcon Laboratories</i>			6445740 Incl.GST :12%	475373.33	6921113.33	
2	UPS			30000 Incl.GST :18%	2100	32100	
3	Rate of one fragmentome hand piece			175000 Incl.GST :12%	12906.25	187906.25	
				6650740	490379.58	7141119.58	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Posterior Segment - Victrectomy system							
Labour	1,20,222.00	1,20,222.00	1,20,222.00	1,20,222.00	1,20,222.00	1,20,222.00	1,20,222.00
Comprehensive	3,18,302.00	3,18,302.00	3,18,302.00	3,18,302.00	3,18,302.00	3,18,302.00	3,18,302.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Posterior Segment - Vitrectomy system

Technical Specification

Equipment: Posterior Segment Vitrectomy System - Table Top Model

1. Cut rate and vacuum should be controllable inversely and simultaneously.
2. Vacuum: should have the facility to generate direct venture/ peristaltic vacuum of up to 650 mmHg.
3. Cutter: should have the ability to drive vertical guillotine vitrectomy cutter to go up to 10000 cuts/minute or more.
4. Should have the facility to allow surgeon to select from 3 different duty cycle option at any given cut rate
5. IOP control: should have the capacity to monitor infusion pressure constantly.
6. Illumination: The system should have dual port Xenon/ solid state illumination.
7. The system should recognize the gauge of illuminator connected
8. MIVS Facility: should have the capacity to support MIVS like 23G, 25G and 27G.
9. Other features: should have vented gas forced infusion capability.
10. Should have silicon oil injection capability by using foot control alone.
11. Should have auto fluid /Air exchange.
12. Should have fully programmable foot switch with the facility to change procedural modes through foot switch.
13. Should have the facility of diathermy.
14. Should have facility for the extrusion of sub-retinal fluid.
15. Should have the facility of voice confirmation.
16. Should have the facility of anterior Phaco emulsification with 4 crystal Hand piece
17. Should have the facility for Anterior Vented Gas forced infusion.
18. Should have facility to use variety of Phaco tips like Kelman ABS and micro tips.
19. Should have irrigation & Aspiration mode
20. Should have the facility of fragmentation with the help of 4 crystal ultrasound hand piece.
21. All accessories, connectors, packs tubing, cutters, electrical connectors etc to make the system work at the user Operation Theatre should be included.
22. Provide suitable table for the machine
23. Should operate from 200 to 240 Vac 50 Hz input supply.
24. Accessories to be provided along with: 23G and 25G Vitrectomy packs including endo illuminator, vitrectomy probe, vitrectomy cassette, and other disposable – 25 Nos.
25. Fragmentation hand piece Fragmentation packs – 1 nos

26. Viscous Fluid Packs – 5 nos
27. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission /FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL.
Copy of the certificate/ test report shall be produced along with the technical bid
28. Suitable UPS to be supplied along with the machine