



Running Contract Details	
Equipment Name	Hematology Slide Stainer
Running Contract Valid Till	09-04-2027
Tender Ref No	KMSCL/EP/T552/1785/2024
Tendered Quantity	5
Supplier Name	M/s Tricore Medicals Systems
GST No	32AAHFT0313N1ZC
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
TC 25/1910(1) Cosmos Building Mele Thampanoor Thiruvananthapuram - 695 001	Contact Person	Praveen K.A
	Phone	0471 2333388
	Mobile No	8606078122
	Email	tricoremedicalsystems@gmail.com

Item-wise Price Details							
#	Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total	
1	Hematology Slide Stainer <i>Model & Make : Stainz Plus / Stainz Serentific</i>			1050200 Incl.GST :18%	73514	1123714	
				1050200	73514	1123714	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Hematology Slide Stainer							
Labour	0.00	0.00	0.00	25,105.93	25,105.93	25,105.93	25,105.93
Comprehensive	0.00	0.00	0.00	59,322.03	59,322.03	59,322.03	59,322.03

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Hematology Slide Stainer

Technical Specification

Equipment: Hematology Slide Stainer

1. [Standalone Automated slide stainer for Haematological Slidestaining.](#)
2. It should base on tradition methodology of working.
3. Minimum staining throughput of 100 slides per hour.
4. Following Romanowsky staining methods should be available. Leshman, Wright-Giemsa, MGG, custom protocol.
5. Able to stain slides prepared from peripheral blood, bone marrow, body fluids etc.
6. It should have facility to connect any two stains at a time and have programmable level of intensities to deliver lighter to darker stained slides at simple change of program.
7. Capacity to load minimum 1 and maximum up to 10 slides in one batch of staining run.
8. Able to dry the slides after staining so as to be ready for microscopy within few seconds of the process
9. Facility to monitor stain levels on screen for inventory management is required.
10. Possibility to load normal glass slides which are regularly uses in clinical laboratories. Notuses any special proprietary slides
11. Stains for use on the system should be readily available in liquid form so as to avoid errors of weighing powder.
12. Have automated self-clean cycle and minimum user maintenance.
13. It should not require any special plumbing
14. Touch-screen user interface will be preferred
15. Operating room temperature: up to 40⁰c
16. Storage temp:60⁰ c
17. Relative humidity up to 90⁰ cnon-condensing.
18. Equipment is CE markeD or ISO CERTIFIED.
19. Should have methanol based staining system