

Running Contract Details	
Equipment Name	Fumigation dispenser
Running Contract Valid Till	15-01-2028
Tender Ref No	KMSCL/EP/T575(R)/50/2025
Tendered Quantity	100
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Fumigation dispensor <i>Model & Make : Aerojet Steriz/ Radiant Enterprises</i>	27900 Incl.GST :18%	1953	29853
		27900	1953	29853
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Fumigation dispensor				
Labour	1,000.00	1,000.00	1,000.00	
Comprehensive	1,700.00	1,700.00	1,700.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Fumigation dispensor

Equipment: Fumigation Dispenser

1. Should be suitable for fumigation in Operation Theatre and other sterile clinical areas.
2. Should have a consistent particle size generation of less than one microns.
3. Should have a maximum through put of 4 meters in a closed room.
4. Should have a vortex type design, non-rotating and non-clogging nozzle assembly.
5. Should have tough and corrosion resistant power head housing.
6. Should have a chemical solution tank of minimum capacity 5 litres.
7. Should have a tough and corrosion resistant tank. The bolts, nuts and washers used inside the tank should be stainless steel / brass/ corrosion resistant material.
8. Should have an easy to clean, detachable and non-corrosive chemical solution tank.
9. Should have primary and secondary filtration for the air intake system.
10. Should be compatible with wide range of disinfectants.
11. Should operate from 200 to 240Vac, 50 Hz input supply.
12. Should have a timer device
13. 200ml solution should sterile 1000CFT area. Certification proof should be attached along with the technical bid.
14. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.