

Running Contract Details	
Equipment Name	Coagulation Analyser
Running Contract Valid Till	31-03-2025
Tender Ref No	KMSCL/EP/T400/1168/2021
Tendered Quantity	50
Supplier Name	M/s Transasia Bio Medicals Ltd
GST No	32AAACT2038C1Z2, 29AAACT2038C1ZP,11AAACT2038C1Z6
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Transasia House 8 Chandivali Studio Road Andheri (East) Mumbai 400 072	Contact Person	Mr.Tapash Patra
	Phone	022-40309063
	Mobile No	9321413273
	Email	t.patra@transasia.co.in,v.sreekumar@transasia.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	SEMI-AUTOMATIC COAGULATION ANALYSER <i>Model & Make : ECL 412/ Erba Lachema S.R.O</i>	121540 Incl.GST :18%	8507.8	130047.8
2	Rate and Pack size of APTT(Pack size 50ml) Rate will be freezed for 3 years	18.44 Incl.GST :0%	1.52	19.96
3	Rate and Pack size of PT(Pack size 50ml) Rate will be freezed for 3 years	17 Incl.GST :0%	1.4	18.4
		121575.44	8510.73	130086.17

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
SEMI-AUTOMATIC COAGULATION ANALYSER							
Labour	1,030.00	1,030.00	1,030.00	1,030.00	1,030.00	1,030.00	1,030.00
Comprehensi	2,060.00	2,060.00	2,060.00	2,060.00	2,060.00	2,060.00	2,060.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
ve							

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :SEMIAUTOMATIC COAGULATION ANALYSER

Semiautomatic Coagulation Analyser

1. The equipment should be capable of analyzing PT and aPTT.
2. Should be an open system with provision to use 3rd party reagents.
3. The cost/ test for PT&APTT should be quoted by the bidder. For the L1 calculation, the total cost of 1000 tests will also be taken and the calculation will automatically be done by the BOQ.
4. Should use optical/mechanical detection principle and there should not be any interference from lipemic /icteric /hemolytic samples.
5. Should handle stat samples.
6. The system should be complete with all incubation steps inbuilt.
7. Should have automatic INR calculation.
8. Should quote available reagent pack sizes and rates for each parameter separately.
9. The bidder shall provide the quantity of reagents required to perform one test and also the reagent requirement as per the format in the offer form. The standard pack should also mention in the offer form. The price bid will not be considered for evaluation if any discrepancy in the quantity mentioned in the offer form and price bid. One standard pack of reagents shall be provided along with equipment by the L1 bidder either in GMC Thiruvananthapuram OR State Public Health Lab, Thiruvananthapuram, for the evaluation of cost per test offered in the tender. If any discrepancy is noticed during the demonstration, then the offer will not be considered and the firm will be blacklisted for a period of 3 years for misrepresentation of facts.
10. Should have a safety certificate from a competent authority CE-IVD/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid
11. Should be semiautomated 4 channel
12. Should be supplied with inbuilt/external printer with necessary accessories
13. Should have result storing facility
14. Manufacturer prescribed calibrator matching with international standards shall be provided along with reagent.
15. Should quote available consumable pack sizes and rates for consumables (Cuvette, Ball vials, and CaCl₂) separately
16. Should also supply pipettes-50/100microlitre along with the machine.

