3 Years

## KERALA MEDICAL SERVICES CORPORATION LTD

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK



Email:ep.kmscl@kerala.gov.in



Warranty period

**Running Contract Details Equipment Name** Negative Pressure Wound Therapy System with Instillation DeviceNegative Pressure Wound Therapy System with Instillation Device Running Contract Valid Till 02-04-2027 Tender Ref No KMSCL/EP/T473(R)/1599/2022 Tendered Quantity 5 Supplier Name M/s ORIOL HEALTHCARE GST No 32AAEFO4715G1ZL Installation & Delivery Period 8 Week(s) Up-time / PM vist 95% & 4 Visits per year

Supplier`s Details							
Address	Contact Details						
46/2251A	Contact Person	SANTHOSH P R					
CHAKKARAPARAMBU VENNALA P. O ERNAKULAM -682028	Phone	9497183707					
	Mobile No	9846120009					
	Email	oriolcochin@gmail.com					

Item-wise Price Details									
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total					
1	Negative Pressure Wound Therapy System with Instillation Device  Model & Make: ULTA/ KCI US IMC	670000.8 Incl.GST :12%	49412.56	719413.36					
2	Size (Volume) 7.5 x 10 x 3.3 cm	7230.72 Incl.GST :12%	533.27	7763.99					
3	Size (Volume) 12.5 x 18 x 3.3 cm	8680 Incl.GST :12%	640.15	9320.15					
4	Size (Volume) 15 x 26 x 3.3 cm	10939.04 Incl.GST :12%	806.75	11745.79					
5	Canister size - 300 ml	2287.04 Incl.GST :12%	168.67	2455.71					
6	Canister size - 500 ml	2891.84 Incl.GST :12%	213.27	3105.11					
7	Canister size - 1000 ml	4305.28 Incl.GST :12%	317.51	4622.79					

Item-wise Price Details										
				706334.7	2 5	52092.19	758426.91			
Annual / Comprehensive Maintenance Charges (Exl.Tax)										
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year			
Negative Pressure Wound Therapy System with Instillation Device										
Labour	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00	25,000.00			
Comprehensi ve	50,000.00	50,000.00	50,000.00	50,000.00	50,000.00	50,000.00	50,000.00			

# **Other terms & conditions**

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- 2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
- 5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

## **Equipment : Negative Pressure Wound Therapy System with Instillation Device**

#### **Equipment: Negative Pressure Wound Therapy System with Instillation Device**

- 1. Supported Therapy NPWT, NPWTI, Open Abdomen Therapy, Peel and Place Therapy
- 2. Size (Volume) needed:
- a. 7.5 x 10 x 3.3 cm
- b. 12.5 x 18 x 3.3 cm
- c. 15 x 26 x 3.3 cm

The rate for the each size should be offered separately in the BOQ and the same shall be taken for evaluation

- 3. User Interface Digital display.
- 4. Battery life 6 hours
- 5. Battery recharge time 8 hours
- 6. Pressure settings NPWT;- 25mm hg to -200 mmhg, Instillation therapy: 50mmHg to 200 mmHg
- 7. Instillation Method Completely automated volumetric solution delivery has one pump for NPWT and one Pump for instillation
- 8. Should have the following Canister size 300 ml, 500 ml, 1000 ml.

The rate for the each size should be offered separately in the BOQ and the same shall be taken for evaluation

- 9. IV pole clamp- Two step locking system
- 10. Disposables:- Silver Dressings, New Specific dressings for NPWT instillation (Cleance, Medium, Large and small) therapy
- 11. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission/ FDA (US)/ ISO certification. Copy of the certificate shall be produced along with the technical bid
- 12. Alarms- Blockage Alarm, Cassette Not Engaged Alarm, Solution Bag/ Bottle Empty Alarm, Fill Assist Inactive Alarm and system error Alarm (00000001 thru 00000009)