



Running Contract Details	
Equipment Name	AMBUBAG VARIOUS TYPES
Running Contract Valid Till	17-06-2026
Tender Ref No	KMSCL/EP/T522/1A1/2024
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Trivandrum Surgicals and Lab Supplies
GST No	32ERJPS6054N1Z5
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
T.C.2/1094(7) SREELAKSHMI PAZHAYA ROAD MEDICAL COLLEGE P.O TVM.695015	Contact Person	GEETHA . S
	Phone	0471 - 2555040
	Mobile No	09447340147
	Email	trivandrumsurgicals@gmail.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ambubag Adult <i>Model & Make : AMBU BAG/FLEXICARE</i>	300	2800 Incl.GST :12%	206.5	3006.5
2	Ambubag pediatric <i>Model & Make : AMBU BAG/FLEXICARE</i>	300	2800 Incl.GST :12%	206.5	3006.5
3	Ambubag infant <i>Model & Make : AMBU BAG/FLEXICARE</i>	300	2800 Incl.GST :12%	206.5	3006.5
			8400	619.5	9019.5

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ambubag Adult

Technical Specification

- 8 Should have a port in the bag to connect oxygen with reservoir bag

II Ambubag (Various types)

1 Ambu Bag – Infant

- 2 Should have silicon rubber bellow to withstand autoclave at 134 deg. C
- 3 Should provide with autoclavable face mask & Oxygen connecting tube.
- 4 Should be supplied with a carry pouch.
- 5 It should have a bag volume of 300 ml. and a variation of ± 100 ml. will be accepted.
- 6 Should have an expiratory resistance of 2.2cms of water
- 7 Should have an inspiratory resistance of 3.3cms of water.
- 8 It should have controlled flow rates and ventilation, and with reduced airway pressure.

III Ambu Bag – Adult

- 1 Should have silicon rubber bellow to withstand autoclave at 134 deg. C
- 2 Should provide with autoclavable face mask & Oxygen connecting tube.
- 3 Should be supplied with a carry pouch.
- 4 It should have a bag volume of 1700 ml. and a variation of ± 100 ml. will be accepted.
- 5 Should have an expiratory resistance of 2.2cms of water
- 6 Should have an inspiratory resistance of 3.3cms of water.
- 7 It should have controlled flow rates and ventilation, and with reduced airway pressure.

- 8 Should have a port in the bag to connect oxygen with reservoir bag

Note : L1 shall be declared separately for each item

Equipment :Ambubag pediatric

Technical Specification

- 8 Should have a port in the bag to connect oxygen with reservoir bag

II Ambubag (Various types)

1 Ambu Bag – Infant

- 2 Should have silicon rubber bellow to withstand autoclave at 134 deg. C
- 3 Should provide with autoclavable face mask & Oxygen connecting tube.
- 4 Should be supplied with a carry pouch.
- 5 It should have a bag volume of 300 ml. and a variation of ± 100 ml. will be accepted.
- 6 Should have an expiratory resistance of 2.2cms of water
- 7 Should have an inspiratory resistance of 3.3cms of water.
- 8 It should have controlled flow rates and ventilation, and with reduced airway pressure.

III Ambu Bag – Adult

- 1 Should have silicon rubber bellow to withstand autoclave at 134 deg. C
- 2 Should provide with autoclavable face mask & Oxygen connecting tube.
- 3 Should be supplied with a carry pouch.
- 4 It should have a bag volume of 1700 ml. and a variation of ± 100 ml. will be accepted.
- 5 Should have an expiratory resistance of 2.2cms of water
- 6 Should have an inspiratory resistance of 3.3cms of water.
- 7 It should have controlled flow rates and ventilation, and with reduced airway pressure.
- 8 Should have a port in the bag to connect oxygen with reservoir bag

Note : L1 shall be declared separately for each item

Equipment :Ambubag infant

Technical Specification

- 8 Should have a port in the bag to connect oxygen with reservoir bag

II Ambubag (Various types)

1 Ambu Bag – Infant

- 2 Should have silicon rubber bellow to withstand autoclave at 134 deg. C
- 3 Should provide with autoclavable face mask & Oxygen connecting tube.
- 4 Should be supplied with a carry pouch.
- 5 It should have a bag volume of 300 ml. and a variation of ± 100 ml. will be accepted.
- 6 Should have an expiratory resistance of 2.2cms of water
- 7 Should have an inspiratory resistance of 3.3cms of water.
- 8 It should have controlled flow rates and ventilation, and with reduced airway pressure.

III Ambu Bag – Adult

- 1 Should have silicon rubber bellow to withstand autoclave at 134 deg. C
- 2 Should provide with autoclavable face mask & Oxygen connecting tube.
- 3 Should be supplied with a carry pouch.
- 4 It should have a bag volume of 1700 ml. and a variation of ± 100 ml. will be accepted.
- 5 Should have an expiratory resistance of 2.2cms of water
- 6 Should have an inspiratory resistance of 3.3cms of water.
- 7 It should have controlled flow rates and ventilation, and with reduced airway pressure.
- 8 Should have a port in the bag to connect oxygen with reservoir bag

Note : L1 shall be declared separately for each item