



Running Contract Details	
Equipment Name	Cytospin
Running Contract Valid Till	01-06-2028
Tender Ref No	KMSCL/EP/T571/1781/2025
Tendered Quantity	5
Supplier Name	M/s LabX
GST No	32BPSPS6759C1ZE
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Opp. SFS Tiara Apartments Near Burma Road Jn. Kumarapuram Medical College P.O Thiruvananthapuram	Contact Person	Mr. Arun
	Phone	471 2447639
	Mobile No	9446004066 ,9446002066
	Email	info@labx.in ,sales@labx.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Cytospin <i>Model & Make : Cytopro / Cyto centrifuge</i>	890900 Incl.GST :18%	62363	953263
2	Cost of filter card	84.25 Incl.GST :18%	5.9	90.15
3	Cost of Reusable funnel	358.53 Incl.GST :18%	25.1	383.63
		891342.78	62393.99	953736.77

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Cytospin							
Labour	15,800.00	15,800.00	15,800.00	15,800.00	15,800.00	15,800.00	15,800.00
Comprehensive	58,550.00	58,550.00	58,550.00	58,550.00	58,550.00	58,550.00	58,550.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cytospin

1. Dimensions (W x D x H): Open: 37.5 x 49 x 59.5 cm, Closed: 37.5 x 49 x 22.5 cm approx.
2. Rated power supply: AC 100 to 240V, 50/60Hz. Power consumption 200 watts maximum
3. Operational temperature range : +10°C to + 40°C
4. Operational relative humidity : 30 to 85%, non-condensing
5. Storage temperature: +0°C to +60°C
6. Storage relative humidity: 20 to 90%, non-condensing
7. Noise Level 50 dBA or lower (When properly balance)
8. Capacity 12/8 Specimen holders, Sample volume 0.3ml to 5ml specimen chambers with atleast 6 mmdiameter cell area & Dual spot-on Slide in one run.
9. Glass Slide sizes: length: 76+0/-1mm, width : 26+0/-1mm, Thickness: 1.1 +0.1/-0.2 mm
10. Rotor operating speed : from 100 rpm to 2000 rpm
11. Program memory: at least 10 programs (from 1 to 99 min)
12. Should have CE IVD certification. Class A device
13. Performance time should be maximum of <5 min per batch with a >95% cell recovery according to CLSI-HS6-P guidelines
14. Rate for filter card and reusable funnel shall be quoted separately in BOQ which will be taken for evaluation