



Running Contract Details	
Equipment Name	Sterile connecting device
Running Contract Valid Till	30-08-2026
Tender Ref No	KMSCL/EP/T532/1029/2024
Tendered Quantity	10
Supplier Name	M/s Haemolab Equipments
GST No	32AAHFH4793A1ZN
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.64/3652 (Old 32/2605) fÃ¢â€šâ€šâ€šâ€šâ€šâ€šâ€š D) - Nilavareth Lane Azad Road - Kaloore Cochin - 682017	Contact Person	Achutha Prakash AJ
	Phone	0484 6999966
	Mobile No	90371 77781, 09387645876
	Email	sales.haemolab@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Sterile connecting device <i>Model & Make : CompoDock/ Fresenius Kabi</i>	1115520 Incl.GST :12%	82269.6	1197789.6
2	Rate of one sterile connection	145.6 Incl.GST :12%	10.74	156.34
		1115665.6	82280.34	1197945.94

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Sterile connecting device							
Labour	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Comprehensive	67,000.00	67,000.00	67,000.00	67,000.00	67,000.00	77,000.00	77,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Sterile connecting device

Technical Specification

Equipment: Sterile Connecting Device

1. Must be usable for quick and safe tubing connections, which maintain a functionally closed system.
2. Information panel should have backlight LCD which provides status and operational prompts.
3. Should have system heat of minimum 300° C (572° F), a temperature high enough to Maintain sterility during cutting and welding without allowing particulates or Chemical residues to form in the tubes.
4. The sealing consumable must be automatically replaced after each use. When the cassette empties or the disposal box is full, a display warning message must be there.
5. The clamping system must be designed to include wet to wet applications.
6. The interconnected left and right clamping mechanism must facilitate connections for efficiency and ease of use.
7. For safety and convenience, If any consumable used it must get ejected into a Built-in disposal box. The disposal box must hold as many as 300 used consumables and an integrated sensor must monitor its volume at all times. These features must help in preventing disposal box overflow.
8. Must work on Ambient temperature: 10°C to 40°C (50°F to 104°F), Relative humidity: 10% to 80%, non-condensing and maximum altitude: 2,000 m (6,562 ft.)
9. Weight must be between 10-23 kg
10. Accessories like Bag supports, fuses (250 V/4.0 A), minimum 100 no's of consumables must be supplied, along with a UPS of suitable capacity.