



Running Contract Details	
Equipment Name	Anatomical Torso Trauma Mannequin
Running Contract Valid Till	11-01-2028
Tender Ref No	KMSCL/EP/T573/1783/2025
Tendered Quantity	100
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	4 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Anatomical Torso Trauma Mannequin <i>Model & Make : PNEUMOTHORAX TRAINER / LAERDAL</i>	198240 Incl.GST :18%	13876.8	212116.8
		198240	13876.8	212116.8

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Anatomical Torso Trauma Mannequin

- 1 Should have upper body torso along with head, neck along with anatomical landmarks allowing needle decompression as per ATLS latest curriculum and ICD (intercostal drain tube) insertion option.
2. Should have Anatomically accurate landmarks aid in site location.
3. Should have bilateral tension pneumothoraxes and decompression portals in each axilla and subclavian region, is specifically designed for training professionals in the practice of chest decompression.
4. Should have Bilateral midaxillary and midclavicular sites available for decompression and ICD insertion.
5. Heimlich Abdominal Thrust Manoeuvre is optional
6. Should be from a reputed manufacturer