

Running Contract Details	
Equipment Name	Trinocular microscope with Camera
Running Contract Valid Till	20-08-2027
Tender Ref No	KMSCL/EP/T541/703C/2025
Tendered Quantity	5
Supplier Name	M/s Unique Scientific
GST No	32AADFU6727F1ZA
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Room No.28/85/16 Ground Floor Capital Tower Patturaikkal Thrissur -680022	Contact Person	Jerin Jose
	Phone	0487-2330038
	Mobile No	9446996490
	Email	unigtcr@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Trinocular microscope with Camera <i>Model & Make : RH-86 UXL/Unilab Microscope</i>	115050 Incl.GST :18%	8053.5	123103.5
2	LED Monitor	38232 Incl.GST :18%	2676.24	40908.24
		153282	10729.74	164011.74

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Trinocular microscope with Camera							
Labour	2,925.00	2,925.00	4,875.00	4,875.00	4,875.00	4,875.00	4,875.00
Comprehensive	4,875.00	4,875.00	9,750.00	9,750.00	9,750.00	9,750.00	9,750.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Trinocular microscope with Camera

Technical Specification

Equipment: Trinocular Microscope with Camera

1. Should have Manual Stage, LED illumination.
2. Abbe condenser (NA 1.25), 7-Position Turret Condenser
3. Should have Focus and centrable
4. Should have iris diaphragm.
5. Should have Trinocular tube eyepieces (10x/20mm, wide-field).
6. Should have Dioptic adjustment on both eyepieces.
7. Should have 5 position nose-piece turret.
8. Should have Brightfield 4x, Phase contrast Objectives 10x,20x,40x,100x Oil
9. Should have Focusing Knobs (Coaxial coarse and fine with stop,Tension control on coarse focus)
10. Suitable mount and HDMI Digital camera should be provided
11. LED Monitor (optional). Rate to be quoted separately in BOQ which will not be taken for evaluation