

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647 Email :ep.kmscl@kerala.gov.in CIN: U24233KL200TSGC021616, PAN : AADCK4029M, GSTIN : 32AADCK4029M1ZK

Running Contract Details				
Equipment Name	VARIOUS MINOR EQUIPMENTS			
Running Contract Valid Till	27-04-2026			
Tender Ref No	KMSCL/EP/T415/143/2021			
Tendered Quantity	Item wise mentioned in below			
Supplier Name	M/s Kerala Surgical Equipment Co			
GST No	32AACFK7295D1ZG			
Installation & Delivery Period	4 Week(s)			
Up-time / PM vist	95% & 0 Visits per year			
Warranty period	1 Years			

Supplier`s Details					
Address	Contact Details				
No.41/1737 A	Contact Person	Mr. Abbas			
MES Centre Near Town Hall Banerji Road	Phone				
Cochin -18	Mobile No	9539012460			
	Email	info@keralasurgicals.com			

Item-wise Price Details							
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total		
1	143 & Weighing machine adult (Make: Doctor Beliram)	625	1410.1 Incl.GST :18%	98.71	1508.81		
2	BP APPARATUS Model & Make : Lifeline	1000	3080 Incl.GST :12%	227.15	3307.15		
	•		4490.1	325.86	4815.96		

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned

above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

## Equipment :143 & Weighing machine adult (Make: Doctor Beliram)

- 1. Should have an accuracy of 500 gms
- 2. Should be dial type having a magnifying lens to see the measurement
- 3. Should measure a maximum weight of 150kgs.
- 4. Should have zero adjustment.

5. Should bear Certification for each machine by department of legal metrology / Weight & Measurement department of India. The tenderer shall submit copy

of model approval issued by Govt. of India, Legal Metrology department along with technical bid.

- 6. Should be Round shape of diameter 300mm (minor variations will be accepted)
- 7. Shall be made of Metal, epoxy powder coated with rust proof parts.

## **Equipment : BP APPARATUS**

- 1. Should be Portable Desktop digital type
- 2. Should be CDSCO approved.
- 3. Should have LED/LCD indication for pressure
- 4. Should have a measuring range from 0 to 300 mmHg.
- 5. Should be provided with latex free adult arm cuffs of size medium & large and paediatric cuff.
- 6. The manometer scale markings and graduations should be permanent and clearly visible
- 7. All plastic parts, if any used should not crack, flake, peel or disintegrate in normal use.
- 8. The inflating latex free bag should be capable of withstanding an internal pressure of 450 mmHg without leaking.
- 9. The inflating bulb should be soft and should not have any joints or ridges.
- 10. The fastening arrangements of the cuff should be of hook and loop type (Velcro).
- 11. The threading and fastening arrangement of the cuff should show no sign of slip or failure when subjected to the maximum test conditions.
- 12. The latex free tubes used should have an internal diameter of  $3 \pm 0.5$  mm and the external diameter should not be less than 8mm.
- 13. The tubes should be fitted with male and female leur connectors.
- 14. The housing case should be of robust design with Aluminium Powder Coated.It should have press to release lock.It should have Metal hinges.The tube should be secured with metal screws and clamps.It sgould have mechanisam to hold the lid in right angles and should prevent accidental dropping.All parts should be replacable dropping.All parts should be replacable in case of breakage.
- 15. Should have Type Approval From Legal Metrology.
- 16. L1 bidder shall be declared individually for each items separately.