



Running Contract Details	
Equipment Name	Examination table
Running Contract Valid Till	15-08-2026
Tender Ref No	KMSCL/EP/T538/1094S/2024
Tendered Quantity	1500
Supplier Name	M/s Carevel Medical Systems Pvt Ltd
GST No	09AABCC6250C1Z1
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
S-35-36 Site- IV Suhibabad Industrial Area Ghaziabad - 201 010	Contact Person	Mr. Gyanendra Rajvanshi
	Phone	0120-4292637
	Mobile No	9599679916
	Email	rkurup@carevelglobal.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Examination table <i>Model & Make : C-3102/Carevel Medical</i>	22774 Incl.GST :18%	1594.18	24368.18
		22774	1594.18	24368.18

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Examination table

Equipment: Examination Table

1. Over all approximate size 1890mm Lx 560 mm W x 840mmH.
2. Fixed upholstered top in two pieces of 56mm width, Rexin shall be of good quality, Mattress shall be PU form of density 50.
3. Thickness of sheets shall be 1.2mm. Thickness of frames shall be 2mm.
4. Head rest adjustable on gas spring
5. Upper section of box with three drawers
6. BP apparatus tray provided near head rest
7. Should provide foot step
8. Pre-treated and powder coated
9. Lower section below the drawers should have storage facility with hinged doors.

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail