

Running Contract Details	
Equipment Name	STERILIZER VARIOUS TYPES
Running Contract Valid Till	11-01-2028
Tender Ref No	KMSCL/EP/T574(R)/115/2025
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Sterilizer small Model & Make : STERILIZER SMALL GIVA 1265/GEETA INDUSTRIES	200	5062.2 Incl.GST :18%	354.35	5416.55
2	Sterilizer medium Model & Make : STERILIZER MEDIUM GIVA 1786 /GEETA INDUSTRIES	200	6478.2 Incl.GST :18%	453.47	6931.67
3	Steriliser large Model & Make : STERILIZER LARGE GIVA 2086 /GEETA INDUSTRIES	200	6842.82 Incl.GST :18%	479	7321.82
			18383.22	1286.83	19670.05

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Sterilizer small

1. Should have 300mm length, 150mm breadth, 125mm height and 1KW heater coil.
2. Should be made up of seamless stainless steel sheet of grade 304.
3. Should have industrial grade heaters with direct power connection
4. Should have sturdy rivetted lid joints with smooth operation.
5. Should have a thermostat, preset at 100 degC maintaining temperature between 98 - 100 degC.
6. Should have indicators and power cord.
7. Should operate on 220/230Vac, 50Hz single phase supply.
8. Should have ISI mark.

Equipment :Sterilizer medium

1. Should have 430mm length, 200mm breadth, 150mm height and 1.5KW heater coil.
2. Should be made up of seamless stainless steel sheet of grade 304.
3. Should have industrial grade heaters with direct power connection
4. Should have sturdy rivetted lid joints with smooth operation.
5. Should have a thermostat, preset at 100 degC maintaining temperature between 98 - 100 degC.
6. Should have indicators and power cord.
7. Should operate on 220/230Vac, 50Hz single phase supply.
8. Should have ISI mark

Equipment :Steriliser large

1. Should have 510mm length, 200mm breadth, 150mm height and 2KW heater coil.
2. Should be made up of seamless stainless steel sheet of grade 304.
3. Should have industrial grade heaters with direct power connection
4. Should have sturdy rivetted lid joints with smooth operation.
5. Should have a thermostat, preset at 100 degC maintaining temperature between 98 - 100 degC.
6. Should have indicators and power cord.
7. Should operate on 220/230Vac, 50Hz single phase supply.
8. Should have ISI mark.