



Running Contract Details	
Equipment Name	Iron Cot SS
Running Contract Valid Till	20-05-2027
Tender Ref No	KMSCL/EP/T550/364B1/2024
Tendered Quantity	1500
Supplier Name	M/s Steel Industries Kerala Ltd
GST No	32AAECS2705F3ZT
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
Post Box No Thuravoor-688532 Alappuzha	Contact Person	
	Phone	
	Mobile No	
	Email	silkgew@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Iron Cot SS <i>Model & Make : Custom/Steel Industrials</i>	24862.6 Incl.GST :18%	1740.38	26602.98
		24862.6	1740.38	26602.98

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Iron Cot SS

1. Overall dimension shall be 2000mm x 900mm x 600mm (LxWxH) tolerance ± 20 mm
2. Frame dimension shall be 1980mm x 900mm (LxW) tolerance ± 20 mm
3. Mattress dimension shall be 1980mmx900mm tolerance ± 20 mm
4. The top sheet shall be 20 gauge(1mm) thickness SS 304 sheet double bend of height minimum reinforced with channels of size minimum 75mm x 30mm flange welded inside the lengthwise frame bends.
5. The bed frame should be provided widthwise 3 nos of SS 304 angle stiffener of size minimum 35x35x2mm and 2 nos angle support size 50x35x2mm one at each head and leg.
6. The leg frames of the cot shall be made of SS 304 tube of diameter 38 mm thick with horizontal bottom tube of diameter 25x1.2 thickness and vertical filling SS 304 tube of diameter 16mm x 1.2mm thickness.
7. The horizontal cross tube should be welded at a height of 270mm from the bottom.
8. The leg frame at the head end should have a height of 1075mm and leg side should be at least 820mm.
9. Angular bracket of 3mm thickness shall be used for fixing the bed frame
10. Sockets should be provided for Mosquito net poles 20mmdiameter and 26mm length
11. The mattress shall be 4" thick, PU foam of 40 density covered with PVC Rexine.
12. Should have provisions to connect IV poles on all four corners of the cot.
13. Should be provided with leg shoes made of PVC bush of height at least 40mm, outside diameter should be 40mm and thickness at least 4 mm.
14. All bolts and nuts used shall be SS 304.
15. The manufacturer should compliant with ISO 9001, 14001, 13485 & OHSAS 18001/OHSAS 45001