KERALA MEDICAL SERVICES CORPORATION LTD



(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647

Email:ep.kmscl@kerala.gov.in

CIN: U24233KL200TSGC021616, PAN: AADCK4029M, GSTIN: 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	Pass Box (static)		
Running Contract Valid Till	23-02-2027		
Tender Ref No	KMSCL/EP/T540/1712B/2024		
Tendered Quantity	15		
Supplier Name	M/s Yorco Sales Pvt Ltd		
GST No	07AAACY0027C1ZX		
Installation & Delivery Period	8 Week(s)		
Up-time / PM vist	95% & 4 Visits per year		
Warranty period	1 Years		

Supplier`s Details						
Address	Contact Details					
Yorco Corporate Tower	Contact Person	Mr Rakes Arora				
Plot No 2 Gazipur - Patparganj	Phone	1146202222				
Delhi - 110096	Mobile No	98911017358				
	Email	sales@yorco.com				

	Item-wise Price Details						
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Pass Box (static) Model & Make : YSL/YSL-207/Yerco sales	87320 Incl.GST :18%	6112.4	93432.4			
		87320	6112.4	93432.4			

Other terms & conditions

- 1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
- $2. \ The \ supplier \ shall \ submit \ performance \ security \ amounting \ to \ 5.00\% \ of \ the \ value \ of \ the \ supply \ order.$
- 3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
- 4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : Pass Box (static)

Technical Specification

Equipment: Pass Box (Static)

- 1. Dimensions :0.6 M x 0.6 M x 0.6 M
- 2. Doors Glass window
- 3. Should have SS handles & SS hinges
- 4. Should have Electromagnetic Interlocking system
- 5. Material of Construction
 - a. Stainless Steel 304
 - b. UV light
 - c. Light indicators for door
 - d. SS Handles & SS hinges
 - e. LED Light
- 6. Availability of Test Report from Central Govt/NABL/ILAC accredited lab to prove conformity to specification