

Running Contract Details	
Equipment Name	MULTIFUNCTIONAL LABOUR COT VARIOUS TYPES
Running Contract Valid Till	26-06-2026
Tender Ref No	KMSCL/EP/T442/64A/2022
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	10 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Multifunctional Labour Cot Type I <i>Model & Make : MEDIK ENTERPRISES</i>	186	65637.5 Incl.GST :18%	4594.63	70232.13
2	Multifunctional Labour Cot Type II <i>Model & Make : MEDIK ENTERPRISES</i>	186	62687.5 Incl.GST :18%	4388.13	67075.63
			128325	8982.75	137307.75

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Multifunctional Labour Cot Type I			
Labour	1,670.00	1,670.00	1,670.00
Comprehensive	4,450.00	4,450.00	4,450.00
Multifunctional Labour Cot Type II			
Labour	1,600.00	1,600.00	1,600.00
Comprehensive	4,250.00	4,250.00	4,250.00

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
ve			

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Multifunctional Labour Cot Type I

Multifunctional Labor Cot Type I

1. Should have a table top with all edges smoothly finished with a foldable protective side rails. The cot shall be completely made up of SS 304.
2. Should have a dimension of L 190 X W 90 X H 80 cm
3. Should be able to hold minimum of 200kg.
4. Should have three sections top. Each section should have separate removable moulded PU form mattress (Velcro should not be used for fixing with table top)
5. Back rest should be adjustable on ratchet.
6. Should have a removable SS collection tray at a distance not exceeding 25 cm from the U-Cut (Table top). The table top section should have slope all around the U-Cut for easy drainage of fluids into the collection tray. The diameter of the collection tray shall be at least 40 cm and depth at least 15 - 20 cm.
7. The U cut should have a diameter of 35 cm and width of this section is 50 cm.
8. Leg section should be foldable or telescopic under the mainframe completely for lithotomy positions. To and fro motion of the leg section should be very smooth
9. Should have side rails for fixing accessories across the length of the table top on both sides. Should have S.S. head rest with moulded mattresses, firm removable side railing steel hand grips, a pair of PU moulded SS adjustable Lithotomy rod with leg crutches and support for foot.
10. Should provide with a detachable saline stand having a provision for fixing four positions (Head end and middle position) on the cot. 2 Nos of IV stand should be provided.
11. Base should be mounted on four heavy duty rubber shoes. At least 4 inches length

Note: Material test certificate for the proof of usage of SS304 should be furnished along with invoice and the manufacturer should also submit undertaking that the tested material is used for the construction of the equipment supplied.

Equipment :Multifunctional Labour Cot Type II

Multifunctional Labor Cot Type II

1. Should have a table top with all edges smoothly finished with a foldable protective side rails. The cot shall be completely made up of SS 304.
2. Should have a dimension of L 180 X W 70 X H 80 cm
3. Should be able to hold minimum of 200kg.
4. Should have three sections top. Each section should have separate removable moulded PU form mattress (Velcro should not be used for fixing with table top
5. Back rest should be adjustable on ratchet.
6. Should have a removable SS collection tray at a distance not exceeding 25 cm from the U-Cut (Table top). The table top section should have slope all around the U-Cut for easy drainage of fluids into the collection tray. The diameter of the collection tray shall be at least 40 cm and depth at least 15 - 20 cm.
7. The U cut should have a diameter of 35 cm and width of this section is 50 cm.
8. Leg section should be foldable or telescopic under the mainframe completely for lithotomy positions. To and fro motion of the leg section should be very smooth
9. Should have side rails for fixing accessories across the length of the table top on both sides. Should have S.S. head rest with moulded mattresses, firm removable side railing steel hand grips, a pair of PU moulded SS adjustable Lithotomy rod with leg crutches and support for foot.
10. Should provide with a detachable saline stand having a provision for fixing four positions (Head end and middle position) on the cot. 2 Nos of IV stand should be provided.
11. Base should be mounted on four heavy duty rubber shoes. At least 4 inches length
12. Note: Material test certificate for the proof of usage of SS304 should be furnished along with invoice and the manufacturer should also submit undertaking that the tested material is used for the construction of the equipment supplied.