



| Running Contract Details | |
|--------------------------------|--|
| Equipment Name | Pacemaker (Temporary) - Single Chamber |
| Running Contract Valid Till | 19-03-2027 |
| Tender Ref No | KMSCL/EP/T544/303A/2024 |
| Tendered Quantity | 10 |
| Supplier Name | M/s Sree Gokulam Healthcare Pvt Ltd |
| GST No | 32AAJCS3502K1ZI |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

| Supplier`s Details | | |
|--|-----------------|-----------------------------|
| Address | Contact Details | |
| 43/2675A Dhanya Complex Sastha Temple Road Ernakulam North PO Kochi- 682018 | Contact Person | Anoop |
| | Phone | |
| | Mobile No | 9995813333 |
| | Email | sales@gokulamhealthcare.com |

| Item-wise Price Details | | | | |
|-------------------------|--|---|------------------------------------|------------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | Pacemaker (Temporary) - Single Chamber <i>Model & Make : Pace 101/Osyka Medical GmbH</i> | 168966 Incl.GST :5% | 13291.99 | 182257.99 |
| | | 168966 | 13291.99 | 182257.99 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | | | | |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
| Pacemaker (Temporary) - Single Chamber | | | | | | | |
| Labour | 5,000.00 | 5,000.00 | 5,000.00 | 5,000.00 | 5,000.00 | 5,000.00 | 5,000.00 |
| Comprehensive | 12,816.00 | 13,460.00 | 14,130.00 | 14,840.00 | 15,580.00 | 16,355.00 | 17,175.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Pacemaker (Temporary) - Single Chamber

Technical Specification

Equipment: Pacemaker (Temporary) Single Chamber

1. Should be a Single Chamber Pacemaker (Temporary) for bradycardia treatment before, during or after a surgery.
2. Stimulation burst and permanent stimulation should be available for high pacing rate.
3. Should be compact & easy-to-operate device, particularly suitable for emergency treatments.
4. Safety features, including automatic lead and battery check.
5. Should have continuous monitoring of the battery voltage / LED indications 7.
6. Should have shock and water-resistant housing.
7. Should have Modes AOO, AAI, VOO, VVI
8. Should have pacing rate 40-180 ppm.
9. Should have fast pacing (Burst rate) of 80-200 ppm.
10. Should have Pulse Amplitude of 0.1-17V
11. Should have sensitivity 1.0-20mV
12. Should have minimum battery backup > 200 hours.
13. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
14. Temporary pacing wires- 5 nos should be supplied along with each unit